

This Page Is Inserted by IFW Operations  
and is not a part of the Official Record

## **BEST AVAILABLE IMAGES**

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images may include (but are not limited to):

- BLACK BORDERS
- TEXT CUT OFF AT TOP, BOTTOM OR SIDES
- FADED TEXT
- ILLEGIBLE TEXT
- SKEWED/SLANTED IMAGES
- COLORED PHOTOS
- BLACK OR VERY BLACK AND WHITE DARK PHOTOS
- GRAY SCALE DOCUMENTS

**IMAGES ARE BEST AVAILABLE COPY.**

As rescanning documents *will not* correct images,  
please do not report the images to the  
Image Problem Mailbox.



**PCT**WORLD INTELLECTUAL PROPERTY ORGANIZATION  
International Bureau

## INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

<b>(51) International Patent Classification <sup>5</sup>:</b> <b>G01N 21/27, G01D 18/00</b>	<b>A1</b>	<b>(11) International Publication Number:</b> <b>WO 94/08225</b> <b>(43) International Publication Date:</b> 14 April 1994 (14.04.94)
<b>(21) International Application Number:</b> PCT/US92/08547 <b>(22) International Filing Date:</b> 7 October 1992 (07.10.92)  <b>(71) Applicant (for all designated States except US):</b> ASHLAND OIL, INC. [US/US]; P.O. Box 391, BL2, Ashland, KY 41114 (US).  <b>(72) Inventor; and</b> <b>(75) Inventor/Applicant (for US only):</b> MAGGARD, Steven, M. [US/US]; 106 Plum Circle, Lake Jackson, TX 77566 (US).  <b>(74) Agents:</b> WILLSON, Richard, C., Jr.; P.O. Box 391, BL2, Ashland, KY 41114 (US) et al.  <b>(81) Designated States:</b> AT, AU, BB, BG, BR, CA, CH, CS, DE, DK, ES, FI, GB, HU, JP, KP, KR, KZ, LK, LU, MG, MN, MW, NL, NO, PL, RO, RU, SD, SE, US, European patent (AT, BE, CH, DE, DK, ES, FR, GB, GR, IE, IT, LU, MC, NL, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, ML, MR, SN, TD, TG).		<b>Published</b> <i>With international search report.</i> <i>With amended claims.</i>
<b>(54) Title:</b> SPECTROSCOPIC INSTRUMENT CALIBRATION  <b>(57) Abstract</b>  A calibrated spectrometer can indirectly determine a physical or chemical property of a sample based upon spectral responses measured by the spectrometer with respect to the particular sample. This invention is directed to a method for calibrating or recalibrating a second spectrometer in light of a first spectrometer, or itself, respectively. The calibration employs a unique selection and manipulation of spectral data obtained from both the first and the second instrument. The recalibration employs a unique selection and manipulation of spectral data from the same first instrument, that is obtained both before and after the need for recalibration arises. Instead of modifying the respective responses of the first and second instrument, or the first instrument before and after the need for calibration arises, this invention modifies the calibration equation of the second, or recalibrated instrument, to yield consistent results to those obtained by the first instrument, or the first instrument before it goes out of calibration. A calibration equation is an equation which transforms spectral data of a particular sample at a variety of wavelengths to a calculated value for a chemical or physical property. Generally, the form of such calibration equations is that of a linear combination of absorbances or mathematical transforms of absorbances measured by the first and second instrument for each sample. Surprisingly, and confirmed statistically, the accuracy of the calibrated or recalibrated instrument is maintained and in some instances improved.		

**FOR THE PURPOSES OF INFORMATION ONLY**

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AT	Austria	FR	France	MR	Mauritania
AU	Australia	GA	Gabon	MW	Malawi
BB	Barbados	GB	United Kingdom	NE	Niger
BE	Belgium	GN	Guinea	NL	Netherlands
BF	Burkina Faso	GR	Greece	NO	Norway
BG	Bulgaria	HU	Hungary	NZ	New Zealand
BJ	Benin	IE	Ireland	PL	Poland
BR	Brazil	IT	Italy	PT	Portugal
BY	Belarus	JP	Japan	RO	Romania
CA	Canada	KP	Democratic People's Republic of Korea	RU	Russian Federation
CF	Central African Republic	KR	Republic of Korea	SD	Sudan
CG	Congo	KZ	Kazakhstan	SE	Sweden
CH	Switzerland	LI	Liechtenstein	SI	Slovenia
CI	Côte d'Ivoire	LK	Sri Lanka	SK	Slovak Republic
CM	Cameroon	LU	Luxembourg	SN	Senegal
CN	China	LV	Latvia	TD	Chad
CS	Czechoslovakia	MC	Monaco	TG	Togo
CZ	Czech Republic	MG	Madagascar	UA	Ukraine
DE	Germany	ML	Mali	US	United States of America
DK	Denmark	MN	Mongolia	UZ	Uzbekistan
ES	Spain			VN	Viet Nam
FI	Finland				

-1-

## SPECTROSCOPIC INSTRUMENT CALIBRATION

### BACKGROUND OF THE INVENTION

#### I. Field of the Invention

This invention relates to spectrophotometers capable of measuring electromagnetic radiation, such as infrared, near-infrared, ultraviolet, visible, Ramon spectra, Rayleigh scattering, and the like. More specifically, this invention relates to calibration techniques useful for predicting chemical or physical properties based on observed spectral data, preferably data obtained by means of a spectrophotometer. Still more preferably, this invention embodies near-infrared calibration equations suitable for making predictions or estimates of physical or chemical properties such as octane.

#### II. Prior Art Summary

In J. Sci. Food Agric., 34(1983), p. 1441-1443, B. G. Osborne directly transfers a calibration equation from one instrument to another.

*"Use of Indirect Multivariate Calibration for Quality Control of Agricultural Products by Near-Infrared Spectroscopy,"* G. Puchwein and A. Eibelhuber (Mikrochim. Acta (Wien), 1986 II, 43-51) teach a computer program for calibrating an instrument. The program picks relevant frequencies for spectral data for part of a set of samples with which to calibrate an instrument and then validates with respect to the remaining samples.

*"Near-Infrared Limitations to Silicon Photodetector Self-Calibration,"* J. M. Palmer (SPIE Vol. 499, Optical Radiation Measurements [1984], p. 7-14) teaches self-calibrating silicon photodiode detectors.

*"A Simple Internal Modulation Technique for the Spectral Calibration of Circular Variable Filter Spectrometers in the Near-Infrared,"* Chandrasekhar, Ashok, Bhatt, and Manian (Infrared Physics, Volume 24, No. 6, 1984, p. 571-572) teach the wavelength calibration of telescopes attached to near-infrared instruments for astronomical studies using the frequency dependent signal of a low-pressure mercury arc lamp.

Mark and Workman (Spectroscopy, 311), 28) developed a computer algorithm to find *isonumeric* wavelengths. Isonumeric wavelengths are calibration

-2-

wavelengths which yield the same absorbance readout on a master NIR and a slave NIR. These wavelengths therefore do not change substantially from one instrument to another.

Heckman, Diffie, and Milhous (Analytica Chemica Acta, 192(1987),197) teach a computer program that transfers calibration equations by computing slope and bias corrections for each wavelength in the equation. This method of slope and bias corrections is, in fact, used in Example 3 of this application. In US 4,761,522 R. D. Rosenthal claims a poly-tetrafluoroethylene standard for near-infrared and processes for making that standard. That standard, however, is not used in calibration transfer from one instrument to another.

*Multivariate Instrument Standardization*, by Wang et al., Anal. Chem. 1991, 63, 2750-2756, discloses four standardization methods for either correcting the primary calibration model for use on a secondary instrument, or correcting the spectra acquired on the second instrument to account for the response differences between instruments. It was discovered that the calibration model built from one batch might not be applicable to another batch. However, using mathematics of multivariate calibration, four different approaches were followed to select the best subset of calibration transfer samples available. Among the methods also discussed was methods for correcting the spectra by a process of linear transfer of one spectral response in its intensity from instrument to that of another. Not disclosed in this reference is the use of a calculated dependent variable corresponding to a physical property of interest being used as the key to establishing coefficient for constants in the second calibration equation.

B. G. Osborne, in an article entitled, *Calibration of Instrument for Near-Infrared Spectroscopy*, (48 Spectroscopy, Vol. 4 No. 4, p. 48-55). Osborne discloses in detail the steps for establishing simple correlations by multiple linear regression and also the use of derivatives of linear terms. Validation of calibration curves are discussed; specifically how particular wavelengths are chosen to establish the terms in the multiple linear regression equations, along with the statistical significance of a particular calibration sample. This method can be utilized in this invention for the purpose of defining the calibration equations of the

-3-

calibrated instrument for a range of values within which useful predictions can be made, i.e., the range of interest.

Still other methods for determining calibration equations are those discussed in *Analytical Chemistry Journal* Vol. 60, pages 1193-1202 (1988) by David M. Haaland and Edward V. Thomas. The principles of inverse least squares, partial least squares, classical least squares, and principal component regression analysis are disclosed.

*Liquid Absorption Standards For Ultraviolet, Visible, and Near-Infrared Spectrophotometry*, R. G. Martinek, *J. Amer. Med. Technol.* July-August, 1978, on p. 210-216. Liquid absorbance standards are shown to have the following desirable characteristics: broad absorption peaks; wide wavelength range; high molar absorptivity; stability in solution; readily definable specifications of purity; and minimal spectral temperature coefficients. The purpose of using such standards is to permit calculation of *absolute* absorbance based upon the observed absorbance and the application of an instrument correction factor.

*The Use of Statistics In Calibrating and Validating Near Infrared On-Line Analyzers*, Bruce Thompson, ISA, 1989 Paper #88-0116, uses near-infrared for both process control and statistical process control.

*"Calibration of Instruments for Near-Infrared Spectroscopy"*, B. G. Osborne (*Spectroscopy*, 4(4), p. 48-50) teaches the need for multiplicative scattering correction for solid samples due to particle-size differences in solids and computer transfer of complex equations. Osborne discusses procedures for developing calibration equations of particular relevance and usefulness in the present invention. The selection of appropriate wavelengths and samples are discussed and relevant articles cited further explaining the techniques employed. The transfer of calibration equations from one instrument to another are disclosed to require, in the case of most modern instruments, only a change in Intercept value, and computerized procedures allow for differences between spectral data generated by different instruments. These computerized procedures are not taught in detail, but are likely to involve transforming the spectral data of one instrument to be consistent with that of another instrument. This is to be distinguished from the

method involved in the present invention, which instead of transforming the spectral data, the relevant coefficients in the calibration equation for the second instrument are changed from those found in the calibration equation of the first instrument in accordance with a unique procedure discussed in this specification.

In the inventor's doctoral dissertation, *A Chemometric Analysis of a Magnetic Water Treatment Device* available from University Microfilms International, Dissertation Information Service, order number 8919928, there is a discussion of linear regression analysis (simple and multiple) for a linear regression model, and data transformations which can transform non-linear types of data into data which is appropriate for linear regression modeling. Definitions are shown which establish the degree of reliability that the value of one variable (the dependent variable) can be estimated on the basis of values for other variables (the independent variables). In general, the methods for determining regression constants (or coefficients) for the calibration equations discussed in the inventor's doctoral thesis are useful in establishing the calibration equation for the first or reference instrument in the present invention.

*A New Approach to Generating Transferable Calibrations for Quantitative Near Infrared Spectroscopy*, H. Mark and J. Workman, *Spectroscopy* Volume 3, No. 11, p. 28-36, shows that instrument variability makes transfer of calibrations between instruments difficult. However, by selecting the wavelengths and corresponding spectral properties for such wavelengths to meet certain criterion, many of the variations between instruments will not change significantly the calibration equation from one instrument for that calibration equation which would work satisfactorily in another instrument. Not disclosed is how to revise one calibration equation from one instrument to be suitable for use in another instrument where there is a significant change in the calibration equation suitable for the second instrument.

*Tobacco constituents to Tilting Filter Instruments*, R. A. Heckman, J. T. Diffie, L. A. Milhous, *Analytica Chimica Acta*, 192(1987) p. 197-203, teach transferring a near-infrared monochromatic calibration for tobacco constituents to a tilting-filter instrument in order to utilize the experimental calibration equation for



-5-

a sensitive but not rugged laboratory instrument to analyze data in a rugged production oriented instrument to predict properties such as moisture, glycerin, propylene glycol, nicotine, and reducing sugar. A program called MTRAN cannot be used without difficulty unless care was exercised to insure that the tilting-filter instrument is properly configured with filters so as to accommodate the wavelengths in the primary or master calibration equation for the laboratory instrument. Consequently, before the primary calibration can be transformed, the monochromatic data files are compressed to simulate a tilting-filter instrument. The resulting data file is used to re-establish the calibration equation. With the foregoing changes, CALTRAN, a program formerly used only for calibration transfer between filter instruments, could be used to effect transfer from a near-infrared monochromatic instrument to a tilting filter instrument.

Typically, the prior art seeks to make changes in the observed spectra (or mathematical transform thereof) produced in one instrument conform to those produced by a calibrated reference instrument. Only after the spectra from an uncalibrated instrument is brought into coincidence with that spectra from a calibrated instrument did one seek to transfer a calibration equation from one instrument to another. Our experience has been that this tends to introduce errors, uncertainty, and numerous numerical difficulties which reduce the ease and precision that is made possible by the invention of this specification.

U.S. 4,963,745 to Maggard entitled *Octane Measuring Process and Device* discloses how to initially establish a calibration equation for a reference instrument. Specifically disclosed is that spectral features, such as second derivatives, of infrared spectral data at methyne band frequencies, in the range of about 1200 to 1236 nanometers (nm), were highly correlatable to octane values for gasoline-like substances. Numerous patents and articles are cited within the text of the specification, as references cited, and as other publications, all of which including those expressly cited are expressly incorporated herein by reference.

An example of a linear calibration equation is one of the form  $Y_c^i = A^i + \sum_j B_j^i X_{jc}^i$  where a  $Y_c^i$  is the predicted value of some physical or chemical property, such as octane values for a gasoline, C, of a set called sample-c. As

-6-

discussed in Example 7, 226 gasoline samples were measured to determine  $A^i$ , and each  $B_{pc}^j$  for  $j = 1$  to 3, to obtain a multiple regression equation suitable for predicting pump octane values of gasoline samples, using a first instrument; instrument-l, the reference instrument. Pump octane is the average of ASTM methods D2699 and D2700, and is commonly known as  $(R+M)/2$  octane. The inferred reproducibility of this test method is a 95% confidence limit of  $\pm 0.71$  octane numbers over the octane range of 84-94 pump octane. From the second instrument, instrument-k, while in transfectance mode, gasoline absorbances at 1196, 1220, and 1236 nanometers (nm) were used to determine the values for the second derivative of an absorbance spectra, the spectral features,  $X_{sp}^k$ , in the linear calibration equation for inferring a pump octane. The standard error of prediction ("SEP") is the statistical parameter that equals the square root of the mean residual variance of unknown samples.

#### SUMMARY OF THE INVENTION

A calibrated spectrometer can indirectly determine a physical or chemical property of a sample based upon spectral responses measured by the spectrometer with respect to the particular sample. This invention is directed to a method for calibrating or recalibrating a first spectrometer in light of a second spectrometer, or itself, respectively. The calibration employs a unique selection and manipulation of spectral data obtained from both the first and the second instrument. The recalibration of the first instrument employs a unique selection and manipulation of spectral data from the same first instrument, that is obtained both before and after the need for recalibration of the first instrument arises. Instead of modifying the respective responses of the first and second instrument, or the first instrument before and after the need for calibration of the first instrument arises, this invention modifies the calibration equation of the second, or recalibrated instrument, to yield consistent results to those obtained by the first instrument before it goes out of calibration. A calibration equation is an equation which transforms spectral data of a particular sample at a variety of wavelengths to a calculated value for a chemical or physical property. Generally, the form of such

-7-

calibration equations is that of a linear combination of absorbances or mathematical transforms of absorbances measured by the first and second instrument for each sample. Surprisingly, and confirmed statistically, the accuracy of the calibrated or recalibrated instrument is maintained and in some instances improved.

Instruments available on the market today take raw spectral data and digitize or form an analog signal representative of it for use in making calculations. As an essential step in such calculations, a calibration equation in conjunction with said signal defines a dependent variable. This dependent variable is either equal to or indicative of a value for some physical or chemical property of the material from which the raw spectral data was measured. The calibration equation appropriate to a particular instrument will vary depending upon changes in the spectrometer that produces the raw spectral data. To calibrate a spectrometer, one defines an appropriate calibration equation that in conjunction with a signal representative of raw spectra measured by the spectrometer for some material will determine for the value of its dependent variable a value either equal to or indicative of some physical or chemical property of the measured material. In practice, changes in such an equation relate directly to changes in constants found in such an equation. This invention is directed in part to how to determine these constants for an uncalibrated instrument by recourse to a calibrated instrument and its calibration equation. Once these constants are determined, then a new or transferred calibration equation has been found that can be used in conjunction with a digitized signal or an analog signal representative of raw spectra measured by the previously uncalibrated instrument.

For example, if one were to determine a calibration equation, calibration equation-2, for a second instrument, instrument-2, by recourse to a calibrated first instrument, instrument-1, and its corresponding calibration equations, equation-1, the steps in any available order for achieving this in accordance with this invention comprise:

- (a) determining from calibration equation-1 a value indicative of each dependent variable-1 corresponding to each member of a calibration

-8-

the steps in any available order for achieving this in accordance with this invention comprise:

- (a) determining from calibration equation-1 a value indicative of each dependent variable-1 corresponding to each member of a calibration set based on spectral data-1 for each member of said calibration set measured by means of instrument-1; and
- (b) determining constants-2 in said calibration equation-2 so that the sum of absolute differences between:
  - (1) said value indicative of each said dependent variable-2 for each member of said calibration set based upon calibration calibration-2 and data-2 corresponding to each member; and
  - (2) that corresponding value of dependent variable-1 calculated for each said member based upon calibration equation-1 and data-1 corresponding to each member;

is substantially minimized.

#### BRIEF DESCRIPTION OF THE DRAWINGS

Figure 1 is a schematic overview of this invention.

Figure 2 is a flow diagram for Example 5.

Figure 3 is a flow diagram for Example 6.

#### DETAILED DISCUSSION OF THE INVENTION

This invention solves the problem of how to utilize the results from calibrating one instrument for calibrating a different instrument. Alternatively, this invention can deal with recalibration of an instrument to insure consistent results after there has been some change in an instrument's spectral response, such as occur due to changes in light sources, for example due to replacing burned out bulbs, or due to new connections used to collect data from different sources, such as changes in fiber optics used to transmit light or electromagnetic radiation from one location to another, due to changes in path length, or due to changes in spectroscopic mode, such as from transmittance to transreflectance.

1  
2  
3  
4 The invention broadly deals with calibrating a new instrument by recourse  
5 to a reference instrument, but also contemplated is a method for continuously  
6 monitoring an instrument to be sure that it has not changed so much so that a  
7 new calibration equation is required to obtain reliable predictions for some physical  
8 or chemical property of a material of interest.

9  
10 The invention contemplates methods for indirectly determining a variety of  
11 properties-p such as cetane, octane, aromatic-naphthene-olefin content,  
12 viscosity-index, molecular weight, and many others, by means of spectroscopic  
13 data. Spectral data broadly is intended to include, but not be limited to, all  
14 intensity versus wavelength spectra that arise when light, broadly intended to cover  
15 a wide range of frequencies such as radio frequency to ultra-violet interacts with  
16 a material. Generally, any spectral features (as defined in this specification)  
17 derivable from spectral data, which can be correlated to predict values for a  
18 chemical or physical property determined directly or indirectly, are within the  
19 intended scope of this invention.  
20

21 Some benefits made available by this invention include:

- 22  
23 1. Only a small number of samples (compared to the number used in  
24 the original calibration of the reference instrument) is required to  
25 determine the calibration equation for the uncalibrated instrument  
26 because the calibration equation of the reference instrument can be  
27 used. Example 7 shows that a small calibration sample consisting  
28 of only four members can be used.  
29  
30 2. The calibration set, sample-c, does not have to contain any unstable  
31 exemplars that may have been included in the original calibration set  
32 of samples and preferably, in the case of mixtures, such as gasoline  
33 or other petroleum fuels, can contain some of the pure molecules or  
34 easily prepared mixtures thereof.

35 The limitation on appropriate members of sample-c in determining a  
36 calibration equation appropriate to an uncalibrated instrument based upon its  
37 spectral responses depends in part upon the statistical treatment chosen for the  
38 data. For multiple linear regressions, the number of members of sample-c must  
39  
40

-10-

be at least one more than the number of wavelengths (independent variables) used in the model, unless a slope and bias correction or other error reducing step is also used. For example, at least four samples must be used to transfer a calibration equation having three independent variables.

We have found that in addition to a slope and bias adjustment, other error reducing steps such as least squares error reduction, generalized least sum, and perhaps other techniques can further reduce any remaining error in the resulting calibration equation-2.

Preferably, for the reference set (sample-c) to be useful for transferring the statistical variance, co-variance, or function derived from variance from a calibrated instrument to an uncalibrated instrument, or to the same instrument at a later date, the sample-c should disclose spectral features which change as the instrument changes. For purposes of this invention, sample-c should be selected so its absorbance shows changes at the wavelengths of interest when changes in instrument performance occur, but should remain constant (within instrumental limits) when the instrument is operated at standard or typical conditions.

A specific application of one solution derived from this invention is the calibration of a near-infrared spectrophotometer by use of the results derived from a fully or partially calibrated reference near-infrared spectrophotometer. Throughout the specification, reference will be made to one or more spectrophotometers. The fully or partially calibrated spectrophotometer or other instrument will be referred to as the "reference instrument" or "reference spectrophotometer", has its calibration equation fully or partially defined. The instrument to be calibrated with respect to the reference will be referred to as the "uncalibrated instrument" or "uncalibrated spectrophotometer". A reference instrument may become uncalibrated or go out of calibration as a result of a variety of changes either within the instrument itself or from connections between the instrument and the source from which it derives its measurements or spectral data.

This invention enables an uncalibrated spectrophotometer to be calibrated by reference to a reference spectrophotometer by means of a far smaller

-11-

1  
2  
3 calibration set of samples (the *calibration* set of samples) than was required to  
4 fully define a calibration equation for the reference spectrophotometer to begin  
5 with. The calibration set of samples is here referred to as *set of sample-c*, or  
6 simply, *sample-c*. To determine a calibration equation for calculating octane,  
7 aromatics, olefins, polyols, hydroxyl number, etc., for a near-infrared instrument,  
8 such as NIR Systems Model 6500, LT Industries Model 1200, Guided Wave  
9 Model 300 Series; Bran-Luebbe Infralyzer; or Infracor; approximately 200 blends  
10 are used. Unfortunately, since gasolines are blends with components having  
11 different volatilities, it is very difficult, if not impossible, to store these samples to  
12 be run on an uncalibrated spectrophotometer at some later date. In other words,  
13 the samples used to calibrate the reference spectrophotometer need not be  
14 available. Not using the same materials used to calibrate the reference  
15 spectrophotometer compounds the difficulty of calibrating an uncalibrated  
16 spectrophotometer because each member of the set of sample-s of the  
17 approximately 200 had to have its octane or knock values determined in  
18 accordance with ASTM D2699 and D2700, or the like. The procedure for  
19 determining knock values are very time consuming and require multiple  
20 determinations on the same sample with different engines before reliable  
21 knock-value results are obtained.

22  
23  
24  
25  
26 Unknown prior to this invention was whether pure gasoline-like components,  
27 e.g. specific hydrocarbon molecules, that are readily reproducible could be  
28 selected for the set of sample-c. We have found in the case of near-infrared  
29 having wavelengths in the range 800 to 2500 nm (nanometers), that a set of  
30 sample-c comprising: benzene; methyl t-butylether; iso-octane; n-decane;  
31 1-octene; 2,3,4-trimethylpentane; n-heptane; cyclohexane; n-hexane; or toluene;  
32 when measured in both a reference instrument and an uncalibrated instrument to  
33 produce spectral data from each instrument can provide sufficient data or  
34 information to calibrate the uncalibrated instrument. A calibrated instrument for  
35 purposes of this specification is one that provides stable and reproducible spectral  
36 responses and has a calibration equation by which statistically significant values  
37 for physical or chemical properties of a sample can be inferred with an accuracy  
38  
39  
40

-12-

equal to or exceeding the accuracy of a direct test method (e.g., ASTM) for determining such chemical or physical property.

Broadly, to transfer a calibration equation for octane, any compounds that are near-infrared active, i.e. absorb near-infrared radiation, in 800 nm to 2,500 nm can be used. Examples of such compounds are: n-paraffins, iso-paraffins, aromatics (such as dichlorobenzene or benzene), alcohols (such as methanol or ethanol), olefins (such as 1-octene), halogenated alkanes (such as trichloroethane), ketones (such as methyl ethyl ketone), ethers (such as methyl-t-butyl ether), esters (such as amylacetate), and naphthenes (such as cyclohexane, and methyl cyclohexane). Throughout this specification and claims, standard error of estimate or prediction are all at a 68% confidence limit, unless expressly stated otherwise. Each is respectively referred to also as "SEE" and "SEP".

The calibration equation for a particular instrument predicts a physical or chemical property based on spectral responses or spectral data. When measuring octane values, the particular spectral responses or mathematical transforms correspond to values of the spectral features that are selected for predicting octane values for a particular sample and are explained in detail in US 4,963,745. The equation is preferably linear, but non-linear equations may also be used.

Spectral features are functionally related to spectral data, in that each spectral feature is uniquely defined by such spectral data. An example of such functional relations are mathematical transforms such as derivatives of first, second, and higher order; fourier analysis yielding frequency components and their contributions; difference spectra; integration by segments; and the like. The only limit on choosing such functional relations or mathematic transforms is that any values derived by applying them to spectral data must be correlatable as independent variables to accurately determine the value for a dependent variable in the calibration equation which corresponds, directly or indirectly, to the value for a physical or chemical property.

In discussing the invention in detail, the following definitions will be helpful. A "function" for purposes of this application is a correspondence that associates



-13-

with each set of values for a set of independent variables, a unique value for a dependent variable. In accordance with this definition, a "calibration equation" of this invention is a function that establishes a correspondence between independent variables having values determined by spectral features, and a dependent variable that corresponds to the value for a chemical or physical property. Any function that determines values for a spectral feature that are both uniquely associated with spectral data and can be used in a *calibration equation* is within the scope of the invention. Spectral data are uniquely defined by the results of measuring spectral responses of a material to electromagnetic radiation. Examples of electromagnetic radiation have been discussed in other portions of this specification. The prior art has established a variety of functions that can be used to derive values for spectral features from spectral data. So long as there exists a calibration equation which contains independent variables equal to the spectral features, defined by a function, there are no limitations as to the particular spectral features or functions for spectral data that determine such spectral features which can be employed with the invention.

"Calibration equation" for purposes of this specification and claims means an equation that predicts directly or indirectly; e.g., is functionally related to, the amount of a chemical or physical property (the dependent variable) based on one or more spectral features (the independent variables) and perhaps some other parameters; e.g., temperature.

A near-infrared regression model for octane measurements in gasoline which contain MTBE was published in a book of proceedings by the American Chemical Society, Division of Fuel Chemistry, in Volume 35, Number 1, Pages 266-275 which was written by Steven M. Maggard and also presented by him at the 199th American Chemical Society National Meeting in Boston, Massachusetts on April 22-27, 1990. Specifically disclosed in this paper and presentation is the use of methyne t-butyl combination bands, including overtones, in models to predict octanes of MTBE-containing gasoline compositions.

"Spectral features," for purposes of this application, means absorbance, transmittance, reflectance (diffuse, specular, total, attenuated total, etc.), emission,

-14-

fluorescence, phosphorescence, or transreflectance, or mathematical functions thereof such as derivatives, ratios, fourier transforms, and differences for ultraviolet, visible, near-infrared, mid-infrared, far-infrared, microwave, or radio frequency electromagnetic radiation.

Throughout this text we have referred to the use of derivatives of spectral absorbance data. Here we will show an example of what we mean when we use the term derivative. The logic used to take a derivative of spectral absorbance data is based on the calculus of finite differences since the data is expressed at regularly spaced intervals and is discontinuous in between the spaced intervals.

Spectral absorbance data such as gathered by a NIRSystems OL-6500 can be used. The OL-6500 has a spectral bandpass of about 9 nm but the spectral data is digitized such that recordings are displayed at 2 nm intervals. In general two things may be varied when the derivative about a point is calculated; the first of these is termed the segment and the second of these is termed the gap. The segment determines the number of points which are used to calculate the average number to the right and left of the point for which a derivative is being taken. The gap on the other hand refers to the number of points about the midpoint (immediately adjacent to the point where the derivative is being taken) which are excluded from the calculation of the adjacent segments (although the total segment length is unaltered). If we refer to the average value of the segment (higher wavelengths) above the point at which the derivative is being taken as A and the average value of the segment below the point where the derivative is taken as B, then according to the calculus of finite differences the first derivative of the point is A minus B.

For example, when the segment is 10 in 2 nanometer increments/segments and the gap is zero, the first derivative of the spectral absorbance ( $d^1A/dw$ ) at 1170 nm is calculated as follows:

$$d^1A/dw_{1170} = \{(A_{1172} + A_{1174} + A_{1176} + A_{1178} + A_{1180})/5\} - \{(A_{1168} + A_{1166} + A_{1164} + A_{1162} + A_{1160})/5\}$$

If the segment is changed to six and the gap is changed to 8 the first derivative is:

-15-

$$d^1A/dw_{1170} = \{(A_{1178} + A_{1176} + A_{1180})/3\} - \{(A_{1164} + A_{1162} + A_{1160})/3\}$$

In a general way, for data occurring at two nanometer intervals the first derivative of spectral absorbance data at wavelength  $x$  is:

$$d^1A/dw_x = \sum_i (A_{(x+(gap/2)+2i)})/(segment/2) \\ - \sum_i (A_{(x-(gap/2)-2i)})/(segment/2)$$

where:

$x$  = wavelength in nanometers, e.g. 1170 and

$i$  = 1 to segment/2

In NIRSystems software package, NSAS (near infrared spectral analysis software, NIRSystems, Inc.), for reasons which are unknown, the above equation is strictly only true for segments of 2, 6, 10, 14, 18 etc., and for gaps of 0, 4, 8, 12, 16, etc. A segment of 0 is treated as a segment of 2. Likewise a segment of 4 is treated as 6, etc. Similarly a gap of 2 is treated as a gap of 0, and a gap of 6 is treated as a gap of 4, etc.

Higher order derivatives may be simplified to expressions involving only first derivative terms and solved by the form of the previous equation. In general when the derivative being sought is of the  $h$ -th order:

$$d^hA/dw_x = d^{h-1}A/dw_{(x+\{(segment+gap)/2\}+1)} \\ - d^{h-1}A/dw_{(x-\{(segment+gap)/2\}-1)}$$

where the gap is equal to gap of the  $h$ -th ordered derivative and the segment used to calculate the  $(h-1)$ -th ordered derivative is the same as that being sought for the  $h$ -th ordered derivative.

-16-

For example, the 4th derivative of 1170 nm at a segment of 10 and a gap of 4 is:

$$\begin{aligned}
 d^4A/dw_{1170} &= d^3A/dw_{1178} - d^3A/dw_{1162} \\
 &= (d^2A/dw_{1186} - d^2A/dw_{1170}) - (d^2A/dw_{1170} - d^2A/dw_{1154}) \\
 &= \{(d^1A/dw_{1194} - d^1A/dw_{1178}) - (d^1A/dw_{1178} - d^1A/dw_{1162})\} \\
 &\quad - \{(d^1A/dw_{1178} - d^1A/dw_{1162}) - (d^1A/dw_{1162} - d^1A/dw_{1146})\} \\
 &= \{ \{ [ (A_{1198} + A_{1200} + \dots + A_{1206})^{/5} - (A_{1190} + A_{1188} + \dots + A_{1182})^{/5} ] \} \\
 &\quad - [ (A_{1182} + A_{1184} + \dots + A_{1190})^{/5} - (A_{1174} + A_{1172} + \dots + A_{1166})^{/5} ] \} \\
 &\quad - [ (A_{1182} + A_{1184} + \dots + A_{1190})^{/5} - (A_{1174} + A_{1172} + \dots + A_{1166})^{/5} ] \} \\
 &\quad - [ (A_{1166} + A_{1168} + \dots + A_{1174})^{/5} - (A_{1158} + A_{1156} + \dots + A_{1150})^{/5} ] \} \\
 &\quad - \{ [ (A_{1182} + A_{1184} + \dots + A_{1190})^{/5} - (A_{1174} + A_{1172} + \dots + A_{1166})^{/5} ] \} \\
 &\quad - [ (A_{1166} + A_{1168} + \dots + A_{1174})^{/5} - (A_{1158} + A_{1156} + \dots + A_{1150})^{/5} ] \} \\
 &\quad - [ (A_{1166} + A_{1168} + \dots + A_{1174})^{/5} - (A_{1158} + A_{1156} + \dots + A_{1150})^{/5} ] \} \\
 &\quad - [ (A_{1150} + A_{1152} + \dots + A_{1158})^{/5} - (A_{1142} + A_{1140} + \dots + A_{1134})^{/5} ] \} \}
 \end{aligned}$$

where all of the lower ordered derivatives are also taken using a segment of 10 and a gap of 4. Note that it is necessary to correct the segment and gap to a valid value before applying one of the generalized formulas (e.g. the fourth

-17-

derivative at a segment of 8 and a gap of 6 is corrected to a segment of 10 and a gap of 4 before applying the formula).

Note segments of 2 nanometers were used, but segments of smaller or larger numbers or fractions of nanometers can be used. The choice preferably depends on the degree of spectral resolution available to the spectrometer.

The Examples of differentials, finite differentials, discussed are but one of many of the types of finite differentials that may be taken. The theory behind all of the methods that are available are substantially the same; however, the specific details involved in the calculation may vary somewhat. It is also to be noted that although finite differentials may be discussed, it is possible that differentials having differences that approach zero, rather than some finite number are also possible. For example, a polynomial power series like a Taylor Series can be established based upon a series of points taken from intensity versus wavelength curves or spectral data.

This invention provides a means for determining a calibration equation for a primary instrument that takes advantage of a substantial amount of work previously expended to obtain a calibration equation for a reference spectrophotometer. The primary instrument may, in fact, be the reference spectrophotometer that has gone out of calibration for reasons that have already been discussed.

To make the discussion easier to understand, the following conventions will be followed. The independent variables appropriate to a calibration equation for a primary instrument will be identified by "-i," or "-1", e.g., independent variable-i or independent variable-1 corresponds to the one or more members of a set of independent variables that are contained within the calibration equation, calibration equation-i, or calibration equation-1, respectively, for the primary instrument, referred to throughout as instrument-i or instrument-1. In most examples, there are three wavelengths used, each wavelength is referred to as wavelength-j, from which three independent variables-i derive values functionally from spectral responses of a sample, c, taken from a set of samples referred to as sample-c. Corresponding to a particular set of independent variables-i, there is a dependent

-18-

variable which will be referred to as dependent variable-i. The value of dependent variable-i will usually be equal to a predicted value of property-p of sample, c. Dependent variable-i corresponds to that dependent variable found in calibration equation-i. The constants in calibration equation-i are referred to as constants-i. There are two different types of constants-i. One corresponds to a slope constant-i, and the other to a bias constant-i. The slope constants-i are coefficients of the independent variables-i. Clearly, there is a slope constant-i for each independent variable-i which is in turn dependent upon the number of wavelengths-j that are used in the calibration equation-i. There is only one bias constant-i which is not a coefficient of any independent variable-i.

When one determines a calibration equation appropriate to instrument-k (also referred to as instrument-2) by recourse to the calibration equation for instrument-i, one essentially determines a new set of calibration equation constants that correspond to constants-k. New constants are found in at least one and preferably two steps, assuming that one already has a calibration equation-i already determined for instrument-i.

To successfully transfer or modify calibration equation-i into calibration equation-k, the SEE or SEP of values for property-p based upon spectral data of instrument-k or instrument-2 by means of calibration equation-k or calibration equation-2 should most preferably be substantially as good as that achieved by instrument-i and calibration equation-i or alternatively be substantially as good as that which would be achieved had instrument-k been calibrated the same way as instrument-i, but in most cases be at least equal or better than the SEE for the primary test method for directly measuring values for property-p used to calibrate instrument-i. The benefit made available by this invention is that the number of samples that must be run on either instrument-i or instrument-k, or both, is significantly less than that number of samples required to obtain a reliable calibration equation-i for instrument-i or to fully calibrate instrument-k in the same way as instrument-i. In other words, whatever the value,  $E(k)$ , in the standard error of prediction for the value for a property-p based upon instrument-k and calibration equation-k will not be substantially greater than the value for the standard error of

prediction resulting from determining values for a property-p by the direct or primary test method initially used to calibrate instrument-i.

One way, but not the only way, to transfer calibration equation-i to instrument-k, is to change the constants of calibration equation-i. These changed constants become constants-k. Calibration equation-i is preferably transformed into calibration equation-k, in two steps. Step Two (below) is optional, but preferable depending upon whether the correction indicated by step two is significant. By significant, is meant the inferred variance in predicted values is not more than about 10% greater than SEE or SEP, whichever is the greater, achieved by calibration equation-k without Step Two.

Step One: Determining calibration equation-k (calibration equation-2) by means of calibration equation-i (calibration equation-1) involves the following sub-steps:

- (a) Determine calibration equation-i by measuring dependent variables-i directly by a primary test method with a standard error of estimate of E, and correlating said dependent variable-i to a set of independent variables-i. The set of independent variables-i may consist of one or more members.
- (b) Obtain respectively with instrument-i and instrument-k spectral data-i and spectral data-k, respectively, for each member of sample-c, which are to be used in step (c) (comprising at least one member, c); functionally determine value-k and value-i, respectively, (equal to each spectral feature-k and -i, respectively) for each member of sample-c corresponding to each independent variable-k and -i, respectively.
- (c) Determine (in accordance with calibration equation-i) a value for said dependent variable-i for each member of sample-c. The value determined for dependent variable-i or a value determined by some function-f using that value determined for dependent variable-i is inserted into calibration equation-k for the value of dependent variable-k in calibration equation-k. Also insert into calibration

-20-

equation-k the values for spectral features-k determined for each member of sample-c. (Recall that spectral features-k are functionally determined by spectral data-k that is determined from measurements made by instrument-k of each member of sample-c. The only unknowns in calibration equation-k will therefore correspond to the constants-k.) The values for constants-k that are to be substituted into calibration equation-k produce a resulting calibration equation-k, wherein the sum of all absolute differences between each estimated value of property-p for sample, c, corresponding to dependent variable-k for sample, c, and that estimated value of property-p for sample, c, corresponding to dependent variable-l for sample, c, is minimized so that values for property-p for sample-u that have values for property-p within the range of interest are inferred based upon the resulting calibration equation-k, with an SEP of not substantially greater than the SEE or SEP of E for determining dependent variable-u directly by some primary test method. Once one has defined the constants-k, that permits one to infer the value of dependent variables-k for members of sample-u, having values for property-p, within the range of interest, within the SEE or SEP of E of the direct or primary test method, then one has obtained a calibration equation-k in light of the former calibration equation-i. An *absolute difference* is the value for a difference between two numbers that results from first squaring and then taking the positive square root. (Recall that instrument-k, the primary instrument, may in fact be the reference instrument that has gone out of calibration, or it may in fact be a totally different instrument).

The value for dependent variable-i that is to be inserted into calibration equation-k for dependent variable-k can be functionally related to dependent variable-i by a function-f such as for example a logarithm (*log*) of any base (e.g., natural or base 10) such as  $Y_c^k = n \cdot \log Y_c^i + m$  or as a linear function,  $Y_c^k = n Y_c^i + m$ , where n and m are constants, or a linear



-21-

combination of these; or in general, any function such as a polynomial or geometric series; provided that over the range of values of interest for  $Y_s^k$ , the function is substantially linear.

The range of values of interest for  $Y_s^k$  is that range of values for a chemical or physical property that calibration equation-k in conjunction with instrument-k have been calibrated to predict with an SEE or SEP of less than that SEE or SEP of the primary test method.

Clearly, if  $Y_c^k = n \bullet Y_c^i + m$  and  $n=1$  and  $m=0$ , then  $Y_o^k = Y_c^i$ . However, for any other values of  $n$  and  $m$ , the corresponding constants will be related to those otherwise determined for  $n=1$  and  $m=0$  as follows:

$$A^k = nA_o^k \quad B_j^k = nB_{jo}^k$$

where  $A_o^k$  = value of constant for  $n=1$  and  $m=0$  and

$B_{jo}^k$  = value of constant for  $n=1$  and  $m=0$  at wavelength-j

In other words, whatever values for  $n$  and  $m$  that are used, the dependent variable-k will be functionally related to dependent variable-i. To determine the value for the corresponding physical or chemical property directly measured, one merely transforms inversely from  $Y_c^k$  to  $Y_o^k$  where for any  $m$  and  $n$ , except  $n=0$ ,  $Y_o^k = (Y_c^k - m)/n$ ;

Instrument-k and its spectral data are useable in this invention, if instrument-k could be calibrated by an identical method to that which could be used to successfully calibrate instrument-i.

Preferably, calibration equation-i and calibration equation-k are linear. In the generalized method discussed above for this invention, calibration equation-i and -k need not either be of the same form or both be linear. By the same form is meant having the same functional dependence on the same number of independent variables with corresponding constants determining a dependent variable.

Preferably, calibration equations for instrument-i and instrument-k have the following linear form:

$$Y_o^i = A^i + \sum_j \{B_j^i X_{jo}^i\} \text{ and } Y_o^k = A^k + \sum_j \{B_j^k X_{jo}^k\}$$

where:

-22-

$Y_s^i$  is the dependent variable-i in calibration equation-i that preferably assumes a value substantially equal to property-p of any member, s, of said sample-s and  $Y_s^k$  is the dependent variable k in calibration equation-k which is functionally related to  $Y_s^i$  by a function-f;

$A^k$  and  $A^i$  are each intercept or bias constants-i and -k in calibration equations-i and -k, respectively;

$X_{js}^i$  and  $X_{js}^k$  are each one of said at least one independent variable-i and -k in calibration equations-i and -k, respectively, having a value equal to at least one said spectral feature-i and -k, respectively of a member, s, at wavelength-j, and;

$B_j^i$  and  $B_j^k$  are slope constants-i and -k, respectively, corresponding to coefficients for each said at least one independent variable-i and -k corresponding to spectral features-i and -k at wavelength-j, respectively.

There may be virtually an infinite variety of ways of substituting values for constants-k that yields a calibration equation-k with an SEP not substantially greater than the SEE or SEP necessarily present in the direct method used at some point to obtain directly or indirectly calibration equation-i. By "substantially greater" is meant 30% greater, but preferably, 20% greater, and still more preferably, 10% greater. Directly determined values of property-p must at least once be obtained by some primary or direct test method, in order to calibrate an instrument by determining its calibration equation for predicting values for a physical or chemical property, property-p, for an array of materials, based upon spectral data for such materials. However, once one has calibrated a first instrument, one can use such instrument to calibrate other instruments without necessarily remeasuring property-p by the primary test method. Instrument-i has a directly-determined calibration equation-i if some or all of the constants-i therein are determined by values directly determined by the primary test method. However, where only values used to determine constants-k arise from values predicted by a previously calibrated instrument, without values from the primary test method for property-p, then instrument-k is indirectly calibrated. In all cases, there will be at least one instrument whose calibration equation is determined

-23-

primarily based upon values for its dependent variable determined by recourse to a primary test method. It is the SEE or SEP of that primary test method which is being referred to above.

However, we have found that the following criteria are particularly useful in determining constants-k assuming that calibration equations-k and -l have a linear form as defined above. One criterion requires that constants-k corresponding respectively to  $B_j^k$  and  $A^k$  minimize to substantially the maximum degree possible the following expression:  $\sum_c (Y_c^k - A^k - \sum_j B_j^k X_{jc}^k)^2$ .

Alternatively, values of  $B_j^k$  and  $A^k$  can be used which substantially satisfy the following two equations simultaneously:  $B_j^k = B_j^l \sum_c X_{jc}^l X_{jc}^k / \sum_c (X_{jc}^h)^2$  and  $A^k = \sum_c (Y_c^k - \sum_j B_j^k X_{jc}^k) / n$ ; where  $n$  = number of members of sample-c.

Finally, we have also found a particularly useful criteria for determining constants-k corresponding to  $B_j^k$  and  $A^k$  is one which involves substantially satisfying the following two equations simultaneously:  $B_j^k = B_j^l \sum_c X_{jc}^l / \sum_c (X_{jc}^h)$  and  $A^k = \sum_c (Y_c^k - \sum_j B_j^k X_{jc}^k) / n$ ; where  $n$  = number of members of sample-c.

Step Two involves a slope and bias correction. The requirements for samples appropriate to a slope and bias correction are discussed in the Examples. Essentially, a slope and bias correction arises from determining values for "m" and "b" which satisfy the following expression:

$$m = [(\sum_c Y_c^a \sum Y_c^h) / n - \sum_c (Y_c^a Y_c^h)] / [\{\sum_c Y_c^k \sum Y_c^h\} / n - \sum_c Y_c^k Y_c^h]$$

$$b = \{\sum_c Y_c^a - m \sum_c Y_c^h\} / n$$

where:

$Y_c^a$  is a directly or indirectly determined value for property-p of a particular member, c, of sample-p that is determined with an inferred standard error of estimate no greater than E of the primary test method; and wherein said value for property-p is within the range of interest;

$n$  is the number of members of sample-p; and

$v$  is the standard error of estimate from using calibration equation-k values to determine dependent variable-k, for sample member, c; wherein  $v$  is not significantly greater than E.

-24-

Substitute said values for m and b into the following expression and an improved calibration equation results for the dependent variable-k for instrument-k, which is as follows:

$$\hat{Y}_c^k = [mA^k + b] + \sum_j [mB_j^k X_{jc}^k].$$

Step Two may in some instances be omitted because the correction suggested or required by applying a slope and bias correction does not result in a significant change in constants-k for calibration equation-k. But it is preferable to carry out the steps involved in a slope and bias correction to determine whether a significant correction is involved or not. We have found in most instances, a slope and bias correction is required.

What is surprising about Step Two, the slope and bias correction, is that even a significant variation in values of constants-k required by or due to the correction does not mean that an error has occurred in carrying out Step One. We have discovered that the error, if any, in Step One is substantially corrected by a slope and bias correction and vice versa. (See Example 1). Further, we have found that the number of samples required to carry out a slope and bias correction, even when added to those required to carry out Step One, are together far fewer in number than was required to determine calibration equation-i for instrument-i. This is totally unexpected and could not have been anticipated. In fact, when a significant correction in constants is indicated by a slope correction, the art (see manual for Near Infrared Spectral Analysis Software published by NIR Systems, Inc. Silver Spring, Maryland, 1989) teaches that one should recalibrate instrument-k by the same method used to initially calibrate instrument-i.

Closer examination of what is happening to the constants-k in calibration equation-k as one applies Steps One and Two is interesting. In Step One, each constant varies substantially independently in order to satisfy the set of numerical criteria discussed above. In Step Two, slope constants all vary together by an amount determined by "m". Intercept constant-k is impacted in two ways, both by "m" and an incremental amount "b". When applied together, a calibration equation-k for instrument-k preferably provides an SEP for dependent variable-k that is almost as good as, and often much better than that achieved by calibration

-25-

equation-i using data measured by Instrument-i or at least an SEP that is equal to or better than E, the value of SEP for the direct test method for determining values used to calibrate instrument-i. This result shown in the examples, is totally unexpected, surprising and valuable.

Also contemplated by this invention is the indirect method for determining a property-p for a sample-u using spectral data-k measured by an instrument-k, capable of determining such spectral data-k for sample-u. Though the art teaches many ways for identifying a calibration equation for an instrument, the art has not provided up till now an easy method for transforming a calibration equation-i appropriate for instrument-i to a calibration equation-k appropriate for instrument-k.

A unique form for calibration equation-k or -i involves the use of adjacent order differentials as spectral features. For example, the calibration equation for instrument-k and instrument-i each have the form:

$$Y_c^k = A^k + \sum_j B_j^k X_{jc}^k + \sum_c C_j^k X_{jc}^{kk}$$

$$Y_c^i = A^i + \sum_j B_j^i X_{jc}^i + \sum_c C_j^i X_{jc}^{ii}$$

Here  $Y_c^i$ , and  $Y_c^k$  are calculated or predicted values for dependent variables based upon the calibration equations for instruments-i and -k, respectively, both for each member, c, from sample-c.

$X_{jc}^i$ , and  $X_{jc}^k$  are second derivatives of observed absorbances for instruments i and k, respectively, both for each member, c, from sample-c at wavelength-j.

$X_{jc}^{ii}$ , and  $X_{jc}^{kk}$  are the third derivative for observed absorbances for instruments-i and -k, respectively, both for sample-c at wavelength-j.

We have discovered that with calibration equations in the form of adjacent orders of derivatives of spectral data, that surprisingly a slope and bias correction brings about a variation in calibration equation constants-k that results in inferred values for property-p of a sample, c, based upon spectral data measured for sample-c by instrument-k that is in excellent agreement with the results found from instrument-i and calibration equation-i.

-26-

**EXAMPLE 1**

A calibration equation for an uncalibrated instrument is obtained based upon the calibration equation established for a reference instrument. Solutions for the following equations determine:  $B_j^k$ 's and  $A^k$  in the linear calibration equation for instrument-k, the uncalibrated instrument, having the form:  $Y_c^k = A^k + \sum_j B_j^k X_{jc}^k$ ; which relies on calibration equation:  $Y_c^i = A^i + \sum_j B_j^i X_{jc}^i$ ; for a calibrated or reference instrument, instrument-i.

Three different mathematical criteria are used independently. Since only three wavelengths are used,  $J = 1$  to 3; and since 25 = total number of members in sample-c,  $c = 1$  to 25. Independent variables,  $X_{jc}^k$ , (where "k" identifies instrument-k) have values for spectral features derived from the spectral response of instrument-k at wavelength-j. Each wavelength-j is within the wavelength range of from 800 to 2500 nm. Here the spectral features are the second derivative of the absorbance determined at three wavelengths-J: 1196, 1220, and 1286 nm.

**Mathematical Criterion I**

$$\begin{aligned} \sum_c (Y_c^i - A^k - \sum_j B_j^k X_{jc}^k) &= 0; \quad \text{Assume } Y_c^i = Y_c^k \\ \sum_c Y_c^i X_{1c}^k - A^k \sum_c X_{1c}^k - B_1^k \sum_c X_{1c}^k X_{1c}^k - B_2^k \sum_c X_{2c}^k X_{1c}^k - B_3^k \sum_c X_{3c}^k X_{1c}^k &= 0; \\ \sum_c Y_c^i X_{2c}^k - A^k \sum_c X_{2c}^k - B_1^k \sum_c X_{1c}^k X_{2c}^k - B_2^k \sum_c X_{2c}^k X_{2c}^k - B_3^k \sum_c X_{3c}^k X_{2c}^k &= 0; \text{ and} \\ \sum_c Y_c^i X_{3c}^k - A^k \sum_c X_{3c}^k - B_1^k \sum_c X_{1c}^k X_{3c}^k - B_2^k \sum_c X_{2c}^k X_{3c}^k - B_3^k \sum_c X_{3c}^k X_{3c}^k &= 0 \end{aligned}$$

Instrument-i is calibrated using multiple linear regression at the second derivative of absorbance at 1196, 1220, and 1236 nm modeling the dependent variable (R+M)/2 octane using 72 gasoline samples encompassing the octane range of 86-93 in an approximate box car distribution. Instrument-i is a NIR Systems 6500 operating in transreflectance mode using fiber optics at a path length of 18mm.

From the regression equation for instrument-i, one obtains values of 25 calibration transfer standards and subsequently assigns these as values of the dependent variable (R+M)/2 octane for the 25 calibration transfer standards. The standards are the same ones seen in Example 2.

-27-

The 25 calibration transfer standards and 10 samples of known octane are run on instrument-k. The results are summarized in Table 1.

Alternatively, we have also determined that it is possible to indirectly derive calibration equation-k from the calibration transferred equation-k, where a least squared error reducing step was used in place of the slope and bias adjustment. When the errors were modeled as the dependent variable-k using the spectral data-k, again as the independent variable, it was possible to sum the resulting constants with the constants from the calibration transferred equation-k and to indirectly derive the least squared solution (mathematical criterion 1).

#### Mathematical Criterion II

$$B_r^k = [B_r^i \sum_c X_{r,c}^i X_{r,c}^k] / [\sum_c (X_{r,c}^k)^2]$$

$$A^k = \sum_c (Y_c^k - \sum_r B_r^k X_{r,c}^k) / n$$

$$\text{Assume } Y_c^i = Y_c^k$$

NOTE:  $c = 1-25$ ; since only three wavelengths 1196, 1220, and 1236 are used; and  $n = 25$ ; since number of members in sample-c is 25.

Instrument-i is calibrated using multiple linear regression at the second derivative of absorbance at 1196, 1220, and 1236 nm modeling the dependent variable  $(R+M)/2$  octane using 72 gasoline samples encompassing the octane range of 86-93 in an approximate box car distribution. Instrument-i is a NIR Systems 6500 operating in transreflectance mode using fiber optics at a path length of 18mm.

From the regression equation for instrument-i, one obtains values of 25 calibration transfer standards and subsequently assigns these as values of the dependent variable  $(R+M)/2$  octane for the 25 calibration transfer standards. The standards are the same ones seen in Example 2.

The 25 calibration transfer standards are run on instrument-k. Regression constants are calculated as follows:

$$B_1^k = B_{1220}^k = [B_{1220}^i \sum_c X_{1220,c}^i X_{1220,c}^k] / \sum_c (X_{1220,c}^k)^2$$

$$A^k = \sum_c [Y_c^i - (B_{1220}^k X_{1220,c}^k) - (B_{1196}^k X_{1196,c}^k) - B_{1236}^k X_{1236,c}^k] / 25$$

-28-

The above constants in the calibration equation predict the value of 10 samples of known octane. There is a severe bias of  $5.314 (R+M)/2$  octane numbers. A computer program establishes that a slope correction of 0.56741 is needed and an intercept correction of 35.646. These adjustments yield the final slope and bias adjusted calibration equation for instrument-k and show a standard error of estimate of 0.184 and bias of 0.0265  $(R+M)/2$  octane numbers on the remaining 62 samples.

### Mathematical Criterion III

$$B_r^k = B_r^i \sum_c X_{p_i}^i / \sum_c X_{p_i}^k$$

$$A^k = \sum_c (Y_c^k - \sum_r B_r^k X_{r_c}^k) / n \quad \text{Assume } Y_c^i = Y_c^k$$

NOTE:  $r = 1-3$ ; since only three wavelengths are used.  $n = 25$ ; since number of members in sample-c is 25. Just as with Criterion II, the above Criterion III are used. The results of all methods are given in Table 1.

To demonstrate the relative accuracy to be expected from an octane calibration equation transferred from one instrument, instrument-i to another, instrument-k, the following is done.

Both instrument-i and instrument-k are each independently calibrated using the same 72 samples. The calibration establishes a linear calibration equation that can be used to predict octanes. A standard error of prediction for each of these equations is established for ten knowns.

The transfer calibration equations for instrument-k is derived from calibration equation-i and one of the following criterion: I, II, or III.

Each transferred equation is slope and bias corrected with ten samples yielding a slope and bias corrected calibration equation.

The results from the above steps are shown in Table 1. What is particularly significant is that the SEP dramatically changes as a result of the slope and bias correction. However, the final SEP, after performing a slope and bias correction, in the case of Criterion I, yields an SEP identical to that of the original calibration equation-i.



-29-

A good estimate of the best SEP one might expect in this instance would be an SEP of 0.146 which is the SEP for instrument-k when fully calibrated using the 72 samples.

**TABLE 1**  
**Calibration Transfer Without Slope & Bias Correction**

	Transfer Calibration Equation-k Using 25 Standards of Table 1			Original Calibration Equation-l (72 Samples)	Independent Calibration of Instrument-k (72 Samples)
	Criterion n I	Criterion II	Criterion III		
$A^k$	74.501	76.041	75.560	76.263	70.918
$B_1^k$ (1220nm)	134.620	206.02	211.90	124.60	131.21
$B_2^k$ (1196nm)	21.139	26.939	26.940	24.407	13.48
$B_3^k$ (1236nm)	69.345	70.681	70.689	65.546	65.251
SEP	0.366	5.383	5.523	0.156	0.146
Bias	-0.288	5.195	5.314	0	0

-30-

## Calibration Transfer With Slope &amp; Bias Correction Using 10 Samples

	Predicted Results of 62 Unknowns Using Transfer Calibration Equation-k			Independent Calibration of Instrument-k (72 Samples)
	Criterion I	Criterion II	Criterion III	
$A^k$	76.013	78.792	78.753	70.918
$B_1^k$ (1220nm)	123.20	116.90	116.86	131.21
$B_2^k$ (1196nm)	19.345	15.285	14.857	13.48
$B_3^k$ (1236nm)	63.460	40.105	38.985	65.251
SEP	0.156	0.184	0.188	0.146
Slope Adjustment (m)	0.91514	.56741	0.55150	1
Bias Adjustment (b)	7.8341	35.646	37.082	0
Bias	0.0191	0.0265	0.0373	0

## EXAMPLE 2

In this example, Mathematical Criterion I of Example 1 (multiple linear regression) is used and the form of the calibration equation is linear, but spectral features equal to second and third adjacent derivatives of spectral data corresponding to each selected frequency are used for the independent variables.

$$Y_c^k = A^k + \sum_j B_j^k X_{jc}^k + \sum_{jc} C_j^k X_{jc}^{lk}$$

$$Y_c^i = A^i + \sum_j B_j^i X_{jc}^i + \sum_{jc} C_j^i X_{jc}^{ii}$$

Here  $Y_c^i$ , and  $Y_c^k$  are calculated or predicted values for dependent variables based upon the calibration equations for instruments-i and -k, respectively, both for each member, c, from sample-c having 8 members.

$X_{jc}^i$ , and  $X_{jc}^k$  are second derivatives of observed absorbances for instruments i and k, respectively, both for each member, c, from sample-c at wavelength-j, where j = 1 to 3.

-31-

$X_{j0}^{ii}$ , and  $X_{j0}^{ik}$  are the third derivative for observed absorbances for instruments-i and -k, respectively, both for sample-c at wavelength-j, where  $j = 1$  to 3.

The three wavelengths are the same as used in Example 1.

Seventy-two gasoline samples are analyzed on instrument-i to obtain their absorbance spectra. In addition, 25 samples of calibration transfer standards are also run. The second and third derivatives of the spectra are then made using segment lengths of 20 and gaps of 0 on NSAS (Near Infrared Spectral Analysis Software, available from NIR Systems, Inc.). The compositions of the 25 calibration transfer standards are shown in Table 7, and are specifically chosen because they encompassed the majority of variance of the 72 gasoline samples at the wavelengths used in the calibration. The 72 gasoline samples have octane in the range of 86.3 - 93.2 (R+M)/2 octane numbers and are distributed among this range in an approximate box car distribution. Each sample is analyzed four times by four different knock engine laboratories (APAL, No. 2 Control Lab, Canton Control Lab, and St. Paul Control Lab) by ASTM methods 2699 and 2700 to provide accurate estimates of (R+M)/2 octane.

The second and third derivative files of the 72 gasoline samples are sent to a mainframe computer, and a multiple linear regression in accordance with Mathematical Criterion I is performed by conventional computer programs using SAS statistics (available from SAS Institute, Inc., Cary, North Carolina) on the second derivative of absorbance at 1196, 1220, and 1236 nm and the third derivative of absorbance at 1196, 1220, and 1236 nm modeling the dependent variable of (R+M)/2 octane. Analysis of the data in this way provides a multiple linear regression model of the form:

$$Y_c' = [(R+M)/2] \text{ octane} = A' + B_1'(X_{1c}') + B_2'(X_{2c}') + B_3'(X_{3c}') + C_1'(X_{1c}'^2) + C_2'(X_{2c}'^2) + C_3'(X_{3c}'^2)$$

The values of the  $B_i$  terms are shown in Table 2. The multiple coefficient of determination ( $R^2$ ), multiple coefficient of correlation (R), F-value and standard error of estimate are shown in Table 2. It can be concluded from Table 2 that the calibration equation is extremely accurate.

-32-

**TABLE 2**  
**Regression Constants Obtained for the**  
**Calibration Equation-I**  
**of the 72 Gasoline Samples**

Variable	Wavelength	Derivative Used	Value
A <sup>1</sup>	Constant	--	71.18951
B <sub>1</sub> <sup>1</sup>	1220 nm	2nd	46.25463
B <sub>2</sub> <sup>1</sup>	1196 nm	2nd	-22.35057
B <sub>3</sub> <sup>1</sup>	1236 nm	2nd	64.61244
C <sub>1</sub> <sup>1</sup>	1220 nm	3rd	-28.34514
C <sub>2</sub> <sup>1</sup>	1196 nm	3rd	-12.83303
C <sub>3</sub> <sup>1</sup>	1236 nm	3rd	-66.41106

**Statistical Results For the Calibration Equation-I**

Multiple Coefficient of Determination	Multiple Coefficient of Correlation	F-Value	Standard Error of Estimate
0.9962	0.9980	2805	0.145

Next the values of the Calibration Transfer Standards are predicted from the calibration equation of instrument-I and the predicted values are used as the dependent variable ((R + M/2) octane) and the absorbance data for the Calibration Transfer Standards on instrument-k are converted to second derivative and third derivative and are used for the values of the independent variables-k to determine relevant constants-k using the classical least squares method or Mathematical Criterion I. The results are given in Table 3.

-33-

**TABLE 3**  
**Regression Constants Obtained on Instrument-k For**  
**the 25 Calibration Transfer Standards**

Variable	Wavelength	Derivative Used	Value
$A^k$	Intercept	--	55.55131
$B_1^k$	1220	2nd	178.49118
$B_2^k$	1196	2nd	-47.22965
$B_3^k$	1236	2nd	130.60840
$C_1^k$	1220	3rd	-71.70553
$C_2^k$	1196	3rd	-50.71056
$C_3^k$	1236	3rd	18.43823

Table 4 shows the statistical data for calibration equation-k.

**TABLE 4**  
**Statistical Data for the Calibration Transfer**  
**to Instrument-K Using the Octanes**  
**Predicted from Instrument-I**

Multiple Coefficient of Determination	Multiple Coefficient of Correlation	F-Value	Standard Error of Estimate
0.9988	0.9994	2544	0.124

Seventy-two gasoline samples were then run on instrument-k and second and third derivative files are obtained as before. The new calibration equation from Table 3 is used to predict the (R+M)/2 octane of the 72 gasoline samples [note that these samples are identical to the original calibration set run on instrument-i] using the same octane as is used in the original calibration of instrument-i. These calculations are done on a PC using a BASIC program. The calculated octanes show a bias of -0.431 and have a standard error of prediction of 0.586. These errors are quite high considering that the original standard error of estimate on instrument-i is only 0.145. Also, it could be seen from the predicted octanes that the error for low octane predictions ((R+M)/2 approximately equal to 86) is quite

-34-

high (i.e., about -0.8) when compared to the higher octane ((R+M)/2 approximately equal to 93) predictions (e.g., much closer to zero, about 0.1). This type of behavior is similar to heteroscedasticity, in that calculated predictions tend to be erroneous by an amount that is a function of estimated octane. For example, lower octanes yielded high errors and higher octanes yielded lower errors. When the above behavior is observed, a slope and bias adjustment is warranted providing that a plot of (Y actual - Y predicted) versus octane approximates a straight line. This ordinarily is the case providing that the original calibration equation is linear or a non-linear equation transformed so as to obtain a linear relationship. From examining the data, it is clear that a slope adjustment is warranted since the relationship approximates a straight line. A PC BASIC program is used to accomplish this and shows that a slope correction for "m" of 0.86406, and a bias correction for "b" of 12.5590 is needed.

$$Y(\text{Predicted}) = \hat{Y}_c^k = [0.86406(A^k + \sum_{i=1}^{72} B_i^k X_{ic}^k + \sum_{i=1}^{72} C_i^k X_{ic}^{lk})] + 12.5590$$

Table 5 shows the final regression constants for instrument-k using the calibration transfer followed by the slope and bias correction.

This equation is found to have a standard error of estimate of 0.190 on instrument-k compared to instrument-i's standard error of 0.145 on the 72 gasoline samples.

**TABLE 5**  
**Final Regression Constants for the Slope**  
**and Bias Corrected Transferred**  
**Calibration for Instrument-i**

Variable	Wavelength	Derivative Used	Value
A <sup>i</sup>	Intercept	—	60.55866
B <sub>1</sub> <sup>i</sup>	1220	2nd	154.22709
B <sub>2</sub> <sup>i</sup>	1196	2nd	-40.80925
B <sub>3</sub> <sup>i</sup>	1236	2nd	112.85349
C <sub>1</sub> <sup>i</sup>	1220	3rd	-61.95788
C <sub>2</sub> <sup>i</sup>	1196	3rd	-43.81697
C <sub>3</sub> <sup>i</sup>	1236	3rd	15.93174

-35-

**EXAMPLE 3**

This example demonstrates the slope and bias correction referred to earlier in this specification (Analytica Chemica Acta, 192(1987),197).

The various terms are as previously defined throughout this specification, the Y-term is a dependent variable calculated or predicted based upon independent variables equal to one or more spectral features identified as X-terms.

Depending upon the size or magnitude of the slope and bias corrections corresponding to m and b; where m is the slope correction and b is the bias correction, one may decide not to modify the calibration equation for the uncalibrated instrument.  $\hat{Y}_o^k$  is the slope-corrected and bias-corrected value predicted for a physical property as shown hereinafter.

In the specific example  $Y_o^k$ , the predicted octane value from calibration equation-k, before the slope and bias correction, and  $Y_c^k$ , the sample measured octane (or octane predicted by a calibrated NIR instrument), are used to determine the slope and bias corrections. In the sample set that is used to do a slope and bias correction, it is necessary that the set:

1. Be representative of the range of interest for a particular property, for example, octane. The sample then should have actual octane values (measured in accordance with ASTM D2699 and D2700) approximately in the range of octane values of interest.
  2. Have the octane measured for each sample, and that this actual octane value be then correlated (in effect plotted) versus the calculated octane value for each sample determined by using the calibration equation to which the slope and bias correction is to be made.
  3. The number, n, of samples used for making the slope and bias correction must be 2 to define both m and b, but in the examples, 10 is usually used.
- $Y_c^k = A^k + \sum B_j^k X_{j_o}^k$  The calibration equation for instrument-k  
 $Y_c^a$  is the directly determined, or actually measured property; here ASTM measured octane.

-36-

4. To find  $m$  and  $b$ , solve for  $m$  and  $b$  which satisfy the following mathematical criteria:

$$\sum_c (Y_c^k Y_c^a) - m \sum_c Y_c^k Y_c^k - b \sum_c Y_c^k = 0$$

$$\sum_c (Y_c^a - m Y_c^k - b) = 0$$

The results are:

$$m = [\sum_c Y_c^a \sum_c Y_c^k] / [n \sum_c (Y_c^a Y_c^k)] / [\sum_c Y_c^k \sum_c Y_c^k] / [n \sum_c Y_c^k Y_c^k]$$

$$b = 1/n \{ \sum_c Y_c^a - m \sum_c Y_c^k \}$$

$$\hat{Y}_c^k = [m A^k + b] + \sum_i [m B_i^k X_{ic}^k]$$

#### EXAMPLE 4

The mathematical treatment is changed from Multiple Linear Regression to Partial Least Squares for calibration transfer to instrument-k from instrument-i.

In this example, instrument-i is a NIR Systems 6500 bench top instrument operating in transfectance mode using fiber optics. The instrument is calibrated as shown in Example 1 using the 72 gasoline samples. The calibration is made as before modeling  $(R+M)/2$  octane using the second derivative of absorbance at 1220, 1196 and 1236 nm as independent variables for the multiple regression calibration equation. The calibration transfer standards are the same 25 as shown in Example 5.

Instrument-k is an uncalibrated NIR Systems OL-6500 operating in transfectance mode using fiber optics. The 25 calibration transfer standards are run on this instrument using the calculated octanes, from instrument-i, as the dependent variable. This invention makes it possible to calibrate instrument-k using a different mathematical approach from that which is used on instrument-i (multiple linear regression). Partial Least Squares (PLS) is chosen as the alternative calibration method for instrument-k. The principles of PLS are well known (Harold Martens and Tormod Naes, Multivariate Calibration, John Wiley & Sons; New York, 1989 [ISBN 471-90979-3]).

PLS is accomplished by modeling the dependent variable of calculated octane for the 25 calibration standards second derivatives of absorbance over the wavelength range of 1190-1240 nm (in 2 nm intervals, i.e., 1190, 1192, 1194 . . . 1240). Cross validation is accomplished using four subset groups. Models are



-37-

developed using one, two, three, four, five, and six latent variables. The four latent variable model seems to provide the best estimate of pump octane for instrument-k (i.e., accounted for the most variance without overfitting the data), and a correlation of 0.9994 with instrument-i's calculated octane values. The four latent variable PLS model is used to predict the values of the first 10 samples of the original 72 sample calibration set. The predicted values show a substantial bias of  $-0.433 (R + M)/2$  octane numbers, and a standard error of 0.468. A slope adjustment of 0.940 and an intercept adjustment of  $5.77 (R + M)/2$  octane numbers is made. The remaining 62 samples are predicted as unknowns. The 62 samples show a bias of 0.0156 octane numbers and a standard error of prediction of 0.153 octane numbers for the unknowns.

It should be noted that a different area of the spectra other than 1190-1240 can be chosen, provided that it is functionally related to octane with a reliability at least as good as that for measuring such octane directly. This also clearly demonstrates that the wavelengths used in calibration equation-1 need not be the same in equation 2 which shows some of the utility of the invention.

#### EXAMPLE 5

This example discloses transferring a calibration equation suitable for MTBE from one refinery to another without MTBE from one knock engine to another, from one instrument after a configuration and lamp change to another instrument, with a substantially decreased error of prediction, by means of a *double* calibration transfer.

This example dramatically demonstrates a utility of the present invention. For example, instrument-k is fully calibrated using only gasolines which contained no MTBE (methyl t-butyl ether). Recently, Congress passed the reformulated fuels act and since instrument-k operates in a refinery that is in a CO (carbon monoxide) non-attainment area, the reformulated fuels act specifies that this refinery must begin adding MTBE (or some other oxygenated feedstock) to its gasoline blends at some future date. Because instrument-k is calibrated without the use of MTBE, it is expected that substantial errors would result if instrument-k's current

-38-

calibration is used on unknown samples which contain MTBE. To keep from completely recalibrating instrument-k, one would want to transfer a calibration equation which can accommodate MTBE as it is added to the refineries gasoline blending stocks at a future date. In this example, instrument-k is a NIR Systems OL-6500 operating in transreflectance mode using fiber optics.

A NIR Systems 6500 is chosen to operate as instrument-i because it was once calibrated using 226 samples, some of which contained MTBE, and can accommodate samples containing 0-15% by volume MTBE.

The problem is somewhat compounded by the fact that the only calibration standards run on instrument-i during its calibration with MTBE were the eight pure compounds (and calculated octanes) shown in Table 6. These eight compounds are not run on instrument-k, so a direct calibration transfer is not possible.

TABLE 6

Isooctane	98.59200
n-decane	42.46400
1-octene	57.33100
2,3,4-trimethylpentane	100.37100
n-heptane	58.10600
Cyclohexane	32.51100
n-hexane	66.06500
Toluene	101.69200

To accomplish the transfer, the eight calibration standards (Table 6) are run on instrument-i again. Since the calibration of instrument-i for the MTBE calibration, instrument-i has been changed from transmission measurements at a path length of 20mm to transreflectance measurements at a path length of 18mm using fiber optics. The lamp and detector gain are changed as well.

Under the new set of conditions, using the eight calibration standards, the calibration for MTBE is transferred from instrument-i in transmission to instrument-i in transreflectance. This equation, without slope and bias or other methods of error correction (e.g., least squares) is then used to calculate predicted octanes for the

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40

-39-

25 calibration standards which are also run on Instrument-i in transfectance and instrument-k in transfectance.

Multiple linear regression on the 25 calibration standards Table 7 on instrument-k using the calculated octane values as dependent variables and the second derivative of absorbance at 1220, 1196, and 1236 nm, yielded instrument-k's new calibration equation which could be used to predict the octanes of gasoline blends containing MTBE.

TABLE 7  
Composition of Calibration Transfer Standards

Sample	Paraffins		Isoparaffins		Aromatics			Naphthenes	Olefins
	Hexane	Heptane	2,3-Dimethyl Butane	Isooctane	Ethylbenzene	S-Butylbenzene	T-Butylbenzene	*1,2-Dimethyl Cyclohexane	1-Octene
1	10%		35%			40%		5%	10%
2	15%		35%			40%		5%	5%
3	10%		40%			40%		5%	5%
4	10%		35%		40%			5%	10%
5	10%		35%		40%			10%	5%
6	15%			30%			40%	10%	5%
7	10%			40%			30%	5%	15%
8	15%			35%			35%	10%	5%
9	10%			45%			30%		15%
10	15%		30%				35%	10%	10%
11	10%		40%			40%			10%
12		15%		40%			30%	5%	10%
13	10%		5%	45%			25%	5%	10%
14		10%	35%		40%			5%	10%
15		15%		35%			35%	10%	5%
16	5%		40%			40%		10%	5%
17	10%		40%			45%		5%	
18	10%		20%	20%		20%	20%	5%	5%
19	5%			45%		40%		5%	5%
20	5%		45%			40%		5%	5%
21	10%			50%			25%	5%	10%
22	15%			50%			25%	5%	5%
23	10%		5%	50%			20%	5%	10%

TABLE 7 (continued)  
Composition of Calibration Transfer Standards

Sample	Paraffins		Isoparaffins		Aromatics			Naphthenes	Olefins
	Hexane	Heptane	2,3-Dimethyl Butane	Isooctane	Ethylbenzene	S-Butylbenzene	T-Butylbenzene	*1,2-Dimethyl Cyclohexane	1-Octene
24	15%		2,3,4-Trimethyl Pentane 30%	10%		15%	15%	5%	10%
25	20%		2,3,4-Trimethyl Pentane 30%	5%		10%	25%	5%	5%

\*Note: 1,2-dimethyl cyclohexane and 1-octene are each a mixture of cis and trans isomers.

-42-

The new equation is slope and bias adjusted on ten gasoline samples. Four of these samples are premium unleaded gasoline blends  $[(R+M)/2 \text{ octane} \approx 92]$  and the remaining six are regular unleaded gasoline blends  $[(R+M)/2 \approx 87]$ . These samples show a bias of -0.919 and a standard error of 0.946  $(R+M)/2$  octane numbers.

From this data, a slope correction of 0.950 and a bias correction of 5.30 is made. The resulting equation shows a bias of 0.0401 and a standard error of prediction of 0.122 on 30 unknown gasoline samples. Some of the pertinent data is summarized below.

SUMMARY TABLE

Parameter	Original Calibration on Instrument-I in transmission	Equation on Instrument-I for 8 calibration transfer standards in transreflectance	Equations on Instrument-k for 25 calibration transfer standards in transreflectance	Final calibration equation n Instrument-k after slope and bias correction
A	87.925	87.318	86.238	87.226
B <sub>1</sub> (1220)	69.864	94.418	102.601	97.471
B <sub>2</sub> (1196)	18.581	23.340	22.488	21.364
B <sub>3</sub> (1236)	27.644	28.058	29.838	28.346
Correlation with $(R+M)/2$ octane	0.992	*	0.997	0.998
Standard Error of Prediction	0.330	*	0.946	0.122
Bias	0	*	-0.919	0.0401

The Summary Table shows a very interesting phenomenon in that the transferred calibration equation is approximately two times more accurate than the original equation. This would appear to stem solely from the fact that the octanes

-43-

used in the slope and bias adjustment for instrument-k are more precisely known than the calibration samples used on instrument-i.

This example also goes against a commonly held opinion of some NIR spectroscopists. These scientists believe that an equation developed for one refinery is not useful for predictions at another. In this example, the MTBE calibration for instrument-i is developed at Ashland's Catlettsburg refinery. Catlettsburg is a very complex refinery by industry standards and has 12 gasoline blending components. Instrument-k, on the other hand, analyzes data at Ashland's St. Paul Park, Minnesota, refinery which has only six gasoline blending components and is relatively much less complex. Moreover, the Catlettsburg and St. Paul Park refineries have almost no overlap in terms of crude slates.

A schematic flow diagram for this example is shown in Figure 2.

#### EXAMPLE 6

This example, schematically outlined in Figure 3, demonstrates that when  $Y_{k,s}$  is a linear function of  $Y_{i,s}$  that the calibration transferred equation is as effective, or approximately as effective, as the equation resulting from the direct calibration transfer of the previous examples;

(i.e., where  $Y_{i,s(\text{predicted})} = Y_{k,s(\text{actual})}$ ). Here we will consider the linear function of  $Y_{i,s}$  where  $Y_{k,s} = 3(Y_{i,s})_{\text{predicted}} + 25$

In a previous example, an  $(R+M)/2$  octane calibration is shown using 226 samples and predicts the  $(R+M)/2$  octanes of 8 calibration standards (isooctane, n-decane, l-octene, 2,3,4-trimethylpentane, n-heptane, cyclohexane, n-hexane and toluene). This calibration is developed on an NIR Systems Model 6500 operating in transmission mode at a path length of 20mm.

Subsequently this instrument was configured for fiber optics using transfectance at a path length of 18mm after the lamp and detector gain setting are changed. The 8 calibration standards are run on the instrument using three times the octane value determined from the transmission measurements plus 25 as the value of the dependent variable. The results of multiple linear regression are:

-44-

$$A_k = 286.951$$

$$B_{k,1220} = 283.252$$

$$B_{k,1196} = 70.020$$

$$B_{k,1236} = 84.185$$

Values are then predicted for ten gasoline samples.

The values of the ten gasoline samples predicted are corrected by three-equivalent methods. Method one involves subtracting 25 from the value predicted for the  $(R+M)/2$  gasoline samples, dividing the result by 3, and performing a slope and bias adjustment. Method 2 involves a slope and bias adjustment using the actually measured octane of the sample. Method three involves using three times the gasoline samples octane plus 25 for the samples used in the slope and bias adjustment and correcting the values predicted for subsequent unknowns by subtracting 25 and dividing by 3. In each of these three methods, a bias of 0.0281 and a standard error of prediction of 0.240  $(R+M)/2$  octane numbers is observed using an unknown set consisting of 62 unknowns.

It should be noted that the final equation for the direct calibration transfer ( $Y_{i,s \text{ predicted}} = Y_{k,s \text{ actual}}$ ) shows a somewhat lower error and bias than by either of the previous methods (bias = 0.0237, SEP = 0.223  $(R+M)/2$  octane numbers).  $Y_{k,s \text{ actual}}$  multiplied by a constant with some other constant added to it yielded a value for the dependent variable for the transfer that had an effect on the error term ( $Y_{k,s \text{ predicted}} - Y_{i,s \text{ actual}}$ ) which is not removed exactly by either the inverse mathematical function or slope and bias adjustment.

However, if we generate the calibration transfer equation (without slope and bias adjustment) and multiply the regression constants ( $A_k$ ,  $B_{k,1220}$ ,  $B_{k,1196}$ , and  $B_{k,1236}$ ) by three and add 25 to  $A_k$ , we can arrive at the exact answer that the direct method ( $Y_{k,s \text{ predicted}} = Y_{i,s \text{ actual}}$ ) yields after using a slope and bias adjustment using the actual  $(R+M)/2$  octanes of the ten samples or by slope and bias adjusting the equation using three times the value of octane plus 25 and subtracting 25 from each predicted octane and dividing by 3.



-45-

**EXAMPLE 7**

A calibration equation for an uncalibrated instrument is obtained based upon the calibration equation established for a reference instrument. Solutions for the following equations determine:  $B_j^k$ 's and  $A^k$  in the linear calibration equation for instrument-k, the uncalibrated instrument, having the form:  $Y_c^k = A^k + \sum_j B_j^k X_{jc}^k$ ; and which relies on or has recourse to the calibration equation:  $Y_c^i = A^i + \sum_j B_j^i X_{jc}^i$ ; for the calibrated or reference instrument, instrument-i.

Mathematical Criterion I, a classical least squares is used. But, since only three wavelengths are used,  $J = 1$  to 3; and since 4 = total number of members in sample-c,  $c = 1$  to 4. Independent variables,  $X_{jc}^k$ , (where "k" identifies instrument-k) have values for spectral features derived from the spectral response of instrument-k at wavelength-j. Each wavelength-j is within the wavelength range of from 800 to 2500 nm. Here the spectral features are the second derivative of the absorbance determined at three wavelengths-J: 1196, 1220, and 1286 nm.

**Mathematical Criterion I**

$$\begin{aligned} \sum_c (Y_c^i - A^k - \sum_j B_j^k X_{jc}^k) &= 0; & \text{Assume } Y_c^i &= Y_c^k \\ \sum_c Y_c^i X_{1c}^k - A^k \sum_c X_{1c}^k - B_1^k \cdot \sum_c X_{1c}^k X_{1c}^k - B_2^k \cdot \sum_c X_{2c}^k X_{1c}^k - B_3^k \cdot \sum_c X_{3c}^k X_{1c}^k &= 0; \\ \sum_c Y_c^i X_{2c}^k - A^k \sum_c X_{2c}^k - B_1^k \cdot \sum_c X_{1c}^k X_{2c}^k - B_2^k \cdot \sum_c X_{2c}^k X_{2c}^k - B_3^k \cdot \sum_c X_{3c}^k X_{2c}^k &= 0; \text{ and} \\ \sum_c Y_c^i X_{3c}^k - A^k \sum_c X_{3c}^k - B_1^k \cdot \sum_c X_{1c}^k X_{3c}^k - B_2^k \cdot \sum_c X_{2c}^k X_{3c}^k - B_3^k \cdot \sum_c X_{3c}^k X_{3c}^k &= 0 \end{aligned}$$

Note in Mathematical Criterion I, the form of the calibration equation for the reference instrument-i need not be the same as that for the uncalibrated instrument-k.

Even after both changing a bulb which would ordinarily change the spectral characteristics of an instrument, and also reconfiguring the instrument for fiber optics operating in transfectance mode at a different path length, it was possible to recalibrate such instrument by means of the steps of this invention.

A calibration is developed on instrument-i for pump octane. The calibration set consists of 226 samples. These samples are analyzed on instrument-i and a multiple linear regression is performed modeling (R+M)/2 octane at the second

-46-

derivative of absorbance at 1220, 1196, and 1236 nm. Instrument-i is operating in transmission mode using a total path length of 20 mm. The regression constants obtained are:

$$\begin{aligned} A^i &= 87.925 \\ B_1^i &= 69.864 \\ B_2^i &= 18.581 \\ B_3^i &= 27.644 \end{aligned}$$

The standard error of estimate for the equation is 0.330. It should be pointed out that these results are from two different knock engine operators running samples on the same knock engines. These results are considered to be good since the ASTM allowable limits for  $(R + M)/2$  are a standard error of estimate of approximately 0.35. These samples were in an approximate box-car distribution from 84-94  $(R + M)/2$  octane.

Based on the multiple regression calibration equation values are predicted for four calibration transfer standards. These standards and their calculated  $(R + M)/2$  octanes are shown below.

Substance	Calculated Octane
Isooctane	98.592
n-decane	42.464
l-octene	57.331
2,3,4-trimethylpentane	100.371

It should also be pointed out that instrument-i is calibrated over a four-month period (i.e., samples were run as they came into the lab) and this is not an optimal way to do a calibration. Also not optimal was the fact that these calibration standards were not run on the instrument promptly after or substantially at the same time as the initial calibration, but not until more than one month after the calibration was completed.

Subsequent to this, the lamp on instrument-i burns out and a new lamp is inserted. The instrument is also configured for fiber optics operating in transfectance mode at a different total path length (~ 18 mm). In addition to the

-47-

configuration mode and path length changes, the detector gain has to be changed to accommodate the new lamp and lower throughput of the fiber optic configuration (as opposed to transmission).

The calibration transfer standards are rerun on instrument-i and a transferred calibration equation is developed modeling calculated octane as a function of the second derivative of absorbance at 1220, 1196, and 1236 nm.

The regression constants obtained were:

$$A^k = 87.717$$

$$B_1^k = 92.471$$

$$B_2^k = 24.079$$

$$B_3^k = 32.054$$

$B_1^k$  is very different from  $B_1^i$  whereas  $A^k$ ,  $B_2^k$  and  $B_3^k$  are within 6 (R+M)/2 octane numbers from  $A^i$ ,  $B_2^i$  and  $B_3^i$ , respectively.

Subsequently, 97 samples are analyzed for (R+M)/2 octane. None of these samples are the same as before. Furthermore they are analyzed by multiple operators on different knock engines from those used before. The calibration transfer equation yields a slight bias of 0.0392 (R+M)/2 octane numbers (note that the bias between two different knock engines is commonly much higher than this) and a standard error of estimate of 0.371 (R+M)/2 octane numbers. A slope and bias adjustment is made using two samples; one at 87.4 octane and one at 92.5 octane, and yields a slope correction of 0.9402 and intercept correction or bias correction of 5.297 and a final equation of:

$$A^k = 87.774$$

$$B_1^k = 86.945$$

$$B_2^k = 22.640$$

$$B_3^k = 30.139$$

The slope and bias adjusted equation has a bias of 0.0 and a standard error of prediction of 0.343.

To demonstrate that this could not be accomplished directly from the original calibration equation from transmission data, the original calibration equation is used to predict the (R+M)/2 octane for each of the 97 gasoline samples. The

-48-

result shows a bias of -0.482 octane. A slope and bias adjustment only using all 97 samples on this original calibration equation shows a bias of 0 and a standard error of estimate of 1.96 (R + M)/2 octane.

Also note that the two point slope and bias adjustment can also be visualized as a generalized least sum error reducing step in the sense that it can be done equivalently using techniques of geometry. For example, defining a line extending through the actual values measured, one can determine constants to multiply the line extending between the values predicted before the adjustment in such a manner as to make the two lines equivalent. Hence this is generalized least sum because the approach does not use squares.

#### MODIFICATIONS

Specific compositions, methods or embodiments discussed are intended to be only illustrative of the invention disclosed by this specification. Variations on these compositions, methods or embodiments are readily apparent to a person of skill in the art based upon the teachings of this specification and are therefore intended to be included as part of the invention disclosed herein.

For example, an equivalent method to those disclosed in the examples, intended to be within the claimed scope of this invention, is determining the bias value for, for example,  $A^k$  in calibration equation-k for instrument-k based upon calibration equation-i for instrument-i, wherein the value for  $A^i$  is assumed to be equal to some arbitrary value, such as zero, or a value determined in connection with  $A^i$  and a function of  $A^i$ ; i.e., a partially calibrated instrument. The "correct" value for  $A^k$ , is that value which when used in a calibration equation-k determines the value for a dependent variable-k, which is substantially equal to the actual value of a physical or chemical property desired. The "correct" value for  $A^k$  can be obtained from employing a slope and bias correction to that equation which resulted from employing an arbitrary value for  $A^i$  or a value determined by some function of  $A^i$ . Another equivalent method like above occurs when the final equation that would otherwise result from a slope and bias equation may be disguised by using a function of either  $m$  or  $b$ , (the terms discussed in Example 6).

1  
2  
3  
4 This will clearly change the calculated or predicted value for the dependent  
5 variable. However, just as was shown in the specification that a function of  $Y$ 's  
6 from calibration equation- $i$  can be used, so also can a function of  $m$  or  $b$  be used  
7 and then the resulting dependent variable transformed back to a value that will be  
8 substantially equal to the actual value for a chemical or physical property of  
9 interest.  
10

11 Still another variation within the scope of this invention involves how values  
12 are assigned to dependent variable- $k$ . These values for dependent variable- $k$ , that  
13 can be used to determine constants- $k$  in calibration equation- $k$  or for performing  
14 a slope and bias correction, can be determined without using a primary test  
15 method but instead by reference to calibration equation- $i$  and spectral data- $i$   
16 obtained from instrument- $i$  or possibly any other calibrated instrument in  
17 conjunction with its corresponding calibration equation which involves substantially  
18 the same range of interest and relevant spectral data.  
19

20 Still another variation of this invention is to use different spectral features in  
21 calibration equation- $i$  from those in calibration equation- $k$ , e.g., different  
22 wavelengths or different mathematical transforms or both; provided such  
23 differences are each correlatable to a physical or chemical property of interest.  
24

25 Reference to documents made in the specification is intended to result in  
26 such patents or literature being expressly incorporated herein by reference,  
27 including any patents or other literature references cited within such documents.  
28

29 In most examples, only three wavelengths were used. The number,  
30 however, can vary depending upon the statistical reliability sought or required. If  
31 the indirect method has an SEP of " $v$ " for some property which is smaller than the  
32 interval of reliability for the direct method for determining such property, then there  
33 is a practical limit to how small it is reasonable to seek to make " $v$ ", and how  
34 many wavelengths should be required.  
35

36 Another modification relates to the spectral features which can be derived  
37 from absorption spectra. In the examples, both second and higher derivatives of  
38 intensities versus wavelengths or absorbances, were used. Clearly, higher  
39 derivatives such as first, second, third, fourth, and fifth or higher are also expressly  
40

-50-

1  
2  
3 contemplated for this invention. Values for a spectral feature can equal simply the  
4 value found for a particular wavelength, or a range of wavelengths around some  
5 reference wavelength; e.g., absorbances. Spectral features are not intended to  
6 suggest that the spectral data is necessarily different in value from a spectral  
7 feature. It is clearly contemplated that it may be more complicated in that a linear  
8 combination of spectral features or mathematical transforms thereof are within the  
9 intended scope of this invention.  
10  
11

12 Also contemplated as a variation of this invention is a linear combination of  
13 still higher adjacent order derivatives, such as third and fourth order derivatives or  
14 higher adjacent order derivatives, of spectral data. See Example 5 for an  
15 example.  
16

17 Further, any set of members or range of numbers cited in the specification  
18 is intended to incorporate expressly herein any member or number falling within  
19 such set or range, respectively, including any subset or range within any set or  
20 range cited.

21 What is claimed is:  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40

-51-

**CLAIMS**

1. A method for calibrating a second instrument by determining constants in a calibration equation, calibration equation-2, for said second instrument by recourse to a first instrument having a calibration equation, calibration equation-1, comprising in any order:

(a) Determining from calibration equation-1 a value indicative of each dependent variable-1 corresponding to each member of a calibration set based on spectral data-1 for each member of said calibration set measured by means of instrument-1; and

(b) Determining constants-2 in said calibration equation-2 so that the sum of absolute differences between:

(1) said value indicative of each said dependent variable-2 for each member of said calibration set based upon calibration equation-2 and data-2 corresponding to each said member; and

(2) that corresponding value of dependent variable-1 calculated for each said member based upon calibration equation-1 and data-1 corresponding to each said member;

is substantially minimized.

2. The method of claim 1, wherein said calibration set has at least one member that yields spectral data-2 when measured by instrument-2 which in conjunction with calibration equation-2 determines a value for said dependent variable-2 for said at least one member which is indicative of a value which is significantly different from that value determined by a direct test method for which said calibration equation-1 and calibration equation-2 were established to model.

-52-

3. The method of Claim 1, wherein calibration equation-1 has the following form:

$$Y^1_s = A^1 + \sum_j (B^1_j \bullet I^1_{js})$$

wherein:

$Y^1_s$  is dependent variable-1 for sample member s;

$A^1$  is an intercept constant-1

$B^1_j$  is a slope constant-1 for every sample member s at wavelength j;

$I^1_{js}$  is an independent variable determined from spectral data-1 corresponding to sample member s at wavelength j;

and said minimization of said absolute differences is achieved by substantially minimizing the following mathematical expression:

$$\sum_s \{ Y^2_s - A^1 - \sum_j (B^1_j \bullet I^2_{js}) \}^2.$$

4. The method of Claim 3, wherein said independent variable is determined from spectral data-1 by taking a mathematical transform of absorbance.

5. The method of Claim 4, wherein said mathematical transform comprises a differential of an absorbance spectra.

6. The method of Claim 1, wherein said calibration equation-1 and calibration equation-2 have the same form shown hereinafter:

$$Y^1_s = A^1 + \sum_j (B^1_j \bullet I^1_{js})$$

wherein:

$Y^1_s$  is dependent variable-1 for sample member s;

$A^1$  is an intercept constant-1

$B^1_j$  is a slope constant-1 for every sample member s at wavelength j;



-53-

$I_{js}^1$  is an independent variable determined from spectral data-1 corresponding to sample member  $s$  at wavelength  $j$ ; and

$$Y_s^2 = A^2 + \sum_j (B_j^2 \cdot I_{js}^2)$$

wherein:

$Y_s^2$  is dependent variable-2 for sample member  $s$ ;

$A^2$  is an intercept constant-2

$B_j^2$  is a slope constant-2 for every sample member  $s$  at wavelength  $j$ ;

$I_{js}^2$  is an independent variable determined from spectral data-2 corresponding to sample member  $s$  at wavelength  $j$ ; and said determining said minimization of said sum of absolute differences comprises setting slope and intercept constants for calibration equation-1 to values established by the following:

$$B_j^1 = B_j^2 \{ \sum_s (I_{js}^2 \cdot I_{js}^1) \} / \{ \sum_s (I_{js}^1 \cdot I_{js}^1) \} \text{ and}$$

$$A^1 = \sum_s \{ Y_s^2 - \sum_j (B_j^1 \cdot I_{js}^1) \} / l$$

where  $l$  is the number of members in said calibration set.

7. The method of Claim 1, wherein said calibration equation-1 and calibration equation-2 have the same form shown hereinafter:

$$Y_s^1 = A^1 + \sum_j (B_j^1 \cdot I_{js}^1)$$

wherein:

$Y_s^1$  is dependent variable-1 for sample member  $s$ ;

$A^1$  is an intercept constant-1

$B_j^1$  is a slope constant-1 for every sample member  $s$  at wavelength  $j$ ;

$I_{js}^1$  is an independent variable determined from spectral data-1 corresponding to sample member  $s$  at wavelength  $j$ ; and

-54-

$$Y_s^2 = A^2 + \sum_j (B_j^2 \bullet I_{js}^2)$$

wherein:

$Y_s^2$  is dependent variable-2 for sample member s;

$A^2$  is an intercept constant-2

$B_j^2$  is a slope constant-2 for every sample member s at wavelength j;

$I_{js}^2$  is an independent variable determined from spectral data-2 corresponding to sample member s at wavelength j; and said determining said minimization of said sum of absolute differences comprises setting slope and intercept constants for calibration equation-1 to values established by the following:

$$B_j^1 = B_j^2 \{ \sum_s I_{js}^2 / \sum I_{js}^1 \} \text{ and}$$

$$A^1 = \sum \{ Y_s^2 - \sum_j B_j^1 \bullet I_{js}^1 \} / n$$

where n is the number of members in said calibration set.

8. The method of Claim X, wherein there is a further improvement of applying a slope and bias correction comprising the step of substituting values for m and b into the following expression:

$$\hat{Y}_s^1 = \{ m \times A(1) + b \} + \sum_j \{ m \times B_j^1 \times I_{js}^1 \}$$

wherein m and b are defined by the following expressions:

$$m = \{ (\sum_c Y_c^a \sum Y_c^k) / n - \sum_c (Y_c^a Y_c^k) \} / \{ (\sum_c Y_c^k \sum Y_c^k) / n - \sum_c Y_c^k Y_c^k \}$$

$$b = 1/n \{ \sum_c Y_c^a - m \sum_c Y_c^k \}$$

where:

Dependent-variable-(ac),  $Y_c^a$ , is selected from the group of values consisting of: a directly determined value for property-p of a member c; a predicted value for property-p of member c, by means of a calibration equation for property-p within

-55-

a range said calibration equation is established to model; wherein each value for said property-p is in a range that said calibration equation-1 was established to model; and a combination thereof;

and

n is the number of members to a set of slope and bias samples that yield calculated values indicative of values within a range that said calibration equation-1 is intended to model.

10. The method of Claim 1, wherein said spectral data-1 comprises absorbances or mathematical transforms thereof.

11. The method of Claim 10, wherein said dependent variable-1 corresponds to a value indicative of octane or cetane and said absorbances or mathematical transforms thereof involve frequencies in the range of 800 to 2500 nanometers.

12. The method of Claim 11, wherein at least one member of said calibration set is selected from compounds active in the range of 800 to 2500 nanometers.

13. The method of Claim 12, wherein said compounds active in the range of 800 to 2500 nanometers includes materials selected from the group consisting of: n-paraffins, iso-paraffins, aromatics, alcohols, olefins, halogenated alkanes, ketones, esters, ethers, and naphthenes.

14. A method for calibrating or recalibrating a first instrument, instrument-k, by reference to a second instrument, instrument-i, comprising in any order:

- (a) obtaining spectral data by means of instrument-k and instrument-i, corresponding respectively to data-k and data-i for each member of a calibration set;

-56-

- (b) determining by means of a calibration equation, calibration equation-i, of said instrument-i, a value, value-i, for a dependent variable, dependent variable-i, corresponding to each member of said calibration set;
- (c) determine a value, dependent variable-f, related by function-f to said value-i for said dependent variable-i, for each member of the calibration set;
- (d) selecting values for said constants-k in calibration equation-k so that the sum of absolute differences between:
- (1) each dependent variable-k calculated based on said calibration-k with said selected constants-k and data-k for each member of said calibration set; and
  - (2) corresponding dependent variable-f for identical members of said calibration set determined by function-f and corresponding value-i for dependent variable-i of each said member;
- is substantially minimized.

- 57 -

## AMENDED CLAIMS

[received by the International Bureau on 10 August 1993 (10.08.93);  
original claims 1-14 replaced by amended claims 1-10 (5 pages)]

1. A method for calibrating a spectral instrument to determine a value of a property of an unknown sample by using constants in a calibration equation-2, for said instrument, by reference to a spectral instrument having a calibration equation-1, said method comprising in any order:

(a) Determining from calibration equation-1 a value of a dependent variable-1 corresponding to each member of a set of calibration transfer standards based on spectral data-1 for each member of said set of calibration transfer standards; and

(b) Determining constants-2 in said calibration equation-2 so as to minimize the sum of absolute differences between:

(1) Each said value of dependent variable-2 for each member of said set of calibration transfer standards using spectral data-2 measured by said instrument; and

(2) That corresponding value of dependent variable-1 determined for each member of said set of calibration transfer standards determined by calibration equation-1.

(c) Determining values of said property in said unknown sample by use of said instrument.

2. A method for calibrating a second spectral instrument to determine a value of a property of an unknown sample by using constants in a calibration equation-2, for said second instrument, by reference to a first spectral instrument having a calibration equation-1, said method comprising in any order:

- 58 -

(a) Determining from calibration equation-1 a value of a dependent variable-1 corresponding to each member of a set of calibration transfer standards based on spectral data-1 for each member of said set of calibration transfer standards measured by means of instrument-1; and

(b) Determining constants-2 in said calibration equation-2 so as to minimize the sum of absolute differences between:

(1) Each said value of dependent variable-2 for each member of said set of calibration transfer standards using spectral data-2 measured by said second instrument; and

(2) That corresponding value of dependent variable-1 determined for each member of said set of calibration transfer standards measured by said first instrument.

(c) Determining values of said property in said unknown sample by use of said second instrument.

3. A method for calibrating or recalibrating a first instrument, instrument-1, by reference to a second instrument, instrument-2, over a range of values for a property of interest comprising in any order:

(a) obtaining spectral data by means of instrument-1 and instrument-2, corresponding respectively to data-1 and data-2 for each member of a set of calibration transfer standards;

(b) determining by means of a calibration equation, calibration equation-1, of said instrument-1, a value for a dependent variable, dependent variable-1, corresponding to each member of said set of calibration transfer standards;

- 59 -

(c) determining a value for dependent variable-2 that is relatable by a function-f to said property of interest, for each member of the set of calibration transfer standards;

(d) selecting values for said constants-1 in calibration equation-1 so that the sum of absolute differences between:

(1) each dependent variable-1 estimated by said calibration-1 with said selected constants-1 and data-1 for each member of said set of calibration transfer standards; and

(2) corresponding dependent variable-2 for identical members of said set of calibration transfer standards,

is substantially minimized.

4. A method according to Claim 1, 2 or 3, wherein said set of calibration transfer standards have values of said dependent variable which are different from the value of said dependent variable which would be obtained by a primary reference test method, and wherein said set of calibration transfer standards have compositions different from the composition of said sample, and wherein the determining of constants-2 is by a method which comprises:

- (a) partial least squares;
- (b) multiple linear regression; and/or
- (c) principal component regression.

5. A method according to Claim 1, 2 or 3, wherein said spectral data-1 and the spectral data-2 comprise infrared spectral data, and wherein said sample comprises liquid, and wherein said sample comprises liquid, and wherein said set of calibration transfer standards comprises at least one composition different from those used to calibrate instrument-1.

- 60 -

6. A method according to Claim 1, wherein said set of calibration transfer standards comprises at least one standard which is a pure compound or mixture of pure compounds.

7. A method according to Claim 1, wherein said set of calibration transfer standards have at least one member that yields spectral data-2 when measured by instrument-2 which in conjunction with calibration equation-2 determines a value for said dependent variable-2 for said at least one member which is indicative of a value which is significantly different from that value determined by a direct test method for which said calibration equation-1 and calibration equation-2 were established to model.

8. A method according to Claim 1, wherein calibration equation-2 has the following form:

$$Y^2_s = A^2 + \sum_j (B^2_j \bullet I^2_{js})$$

wherein:

$Y^1_s$  is dependent variable-1 for sample member s, based upon calibration equation-1 for instrument-1;

$Y^2_s$  is dependent variable-2 for sample member s;

$A^2$  is an intercept constant-2

$B^2_j$  is a slope constant-2 for every sample member s at wavelength j;

$I^2_{js}$  is an independent variable determined from spectral data-2 corresponding to sample member s at wavelength j;

and said minimization of said absolute differences is achieved by substantially minimizing the following mathematical expression:

$\sum_s \{Y^1_s - A^2 - \sum_j (B^2_j \bullet I^2_{js})\}^2$ ; and wherein said independent variable is determined from spectral data-2 by taking a mathematical transform of absorbance, and wherein said mathematical transform comprises a differential of an absorbance spectra.



9. A method according to any preceding claim, wherein there is a further improvement of applying a slope and bias correction, whereby the accuracy of dependent variable,  $Y^1_{,}$ , is increased to  $\hat{Y}^1_{,}$ , comprising the step of substituting values for m and b into the following expression:

$$\hat{Y}^1_{,} = \{ m \times A(1) + b \} + \sum_j \{ m \times B^1_j \times I^1_{,j} \}$$

wherein m and b are defined by the following expressions:

$$m = \{ (\sum_c Y_c^a \sum Y_c^k) / n - \sum_c (Y_c^a Y_c^k) \} / \{ \{ \sum_c Y_c^k \sum Y_c^k \} / n - \sum_c Y_c^k Y_c^k \}$$

$$b = 1/n \{ \sum_c Y_c^a - m \sum_c Y_c^k \}$$

where:

Dependent-variable-(ac),  $Y^a_{,}$ , is selected from the group of values consisting of: a directly determined value for property-p of a member c; a predicted value for property-p of member c, by means of a calibration equation for property-p within a range said calibration equation is established to model; wherein each value for said property-p is in a range that said calibration equation-1 was established to model; and a combination thereof;


and

n is the number of members to a set of slope and bias samples that yield calculated values indicative of values within a range that said calibration equation-1 is intended to model.

10. A method according to Claim 1 through 8, wherein there is a further improvement comprising an error reducing step.

1/3

FIG. 1



## INTERNATIONAL SEARCH REPORT


International Application No

PCT/US 92/08547

<b>I. CLASSIFICATION OF SUBJECT MATTER</b> (If several classification symbols apply, indicate all) <sup>6</sup>		
According to International Patent Classification (IPC) or to both National Classification and IPC		
Int.Cl. 5 G01N21/27; G01D18/00		
<b>II. FIELDS SEARCHED</b>		
Minimum Documentation Searched <sup>7</sup>		
Classification System	Classification Symbols	
Int.Cl. 5	G01N ; G01D	
Documentation Searched other than Minimum Documentation to the Extent that such Documents are Included in the Fields Searched <sup>8</sup>		
<b>III. DOCUMENTS CONSIDERED TO BE RELEVANT<sup>9</sup></b>		
Category <sup>*</sup>	Citation of Document, <sup>11</sup> with indication, where appropriate, of the relevant passages <sup>12</sup>	Relevant to Claim No. <sup>13</sup>
A	ANALYTICAL CHEMISTRY vol. 63, no. 23, 1 December 1991, pages 2750 - 2756 WANG ET AL. 'MULTIVARIATE INSTRUMENT STANDARDIZATION' cited in the application see page 2750 - page 2753 ---	1, 14
A	APPLIED SPECTROSCOPY vol. 46, no. 5, 1 May 1992, pages 764 - 771 WANG ET AL. 'CALIBRATION TRANSFER, ETC.' see page 764 - page 765 ---	1, 14
A	ANALYTICAL CHEMISTRY vol. 64, no. 5, 1 March 1992, pages 562 - 564 WANG ET AL. 'IMPROVEMENT TO MULTIVARIATE CALIBRATION, ETC.' ---	1, 14
-/--		
<p><sup>*</sup> Special categories of cited documents: <sup>10</sup></p> <p>"A" document defining the general state of the art which is not considered to be of particular relevance</p> <p>"E" earlier document but published on or after the international filing date</p> <p>"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)</p> <p>"O" document referring to an oral disclosure, use, exhibition or other means</p> <p>"P" document published prior to the international filing date but later than the priority date claimed</p> <p>"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention</p> <p>"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step</p> <p>"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art</p> <p>"A" document member of the same patent family</p>		
<b>IV. CERTIFICATION</b>		
Date of the Actual Completion of the International Search	Date of Mailing of this International Search Report	
23 JUNE 1993	02.07.93	
International Searching Authority	Signature of Authorized Officer	
EUROPEAN PATENT OFFICE	BOEHM C.E.	

III. DOCUMENTS CONSIDERED TO BE RELEVANT (CONTINUED FROM THE SECOND SHEET)		
Category *	Citation of Document, with indication, where appropriate, of the relevant passages	Relevant to Claim No.
A	<p>ANALYTICAL CHEMISTRY vol. 63, no. 20, 15 October 1991, pages 2329 - 2338 ADHIHETTY ETAL. 'ACHIEVING TRANSFERABLE MULTIVARIATE CALIBRATION, ETC.' see page 2329 - page 2333 ---</p>	1, 14
A	<p>SPECTROSCOPY vol. 4, no. 4, 1 May 1989, pages 48 - 52 B.G.OSBORNE 'CALIBRATION OF INSTRUMENTS FOR NEAR INFRARED SPECTROSCOPY' cited in the application ---</p>	1, 14
A	<p>SPECTROSCOPY vol. 3, no. 11, 1 December 1988, pages 28 - 37 MARK 'A NEW APPROACH TO GENERATING TRANSFERABLE CALIBRATIONS, ETC.' cited in the application see page 28 - page 31 ---</p>	1, 14
A	<p>US, A, 4 864 842 (REGIMAND) 12 September 1989 see column 4 - column 7; figures 6, 7 -----</p>	1, 14

2 / 3

**EXAMPLE 5  
FLOW DIAGRAM**

*Instrument-i in Caltettsburg Refinery  
Instrument-k in St. Paul Refinery*

**Figure 2**

3 / 3

# EXAMPLE 6 Flow Diagram




Figure 3

US 9208547  
SA 65980

**23/06/93**

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US-A-4864842	12-09-89	AU-B- 607081	21-02-91
		AU-A- 3905089	01-02-90
		EP-A- 0353118	31-01-90
		JP-A- 2075943	15-03-90
		US-E- RE34070	22-09-92
-----			

**THIS PAGE BLANK (USPTO)**